

Radiation Dose Monitoring in Pediatrics:

Experience at Boston Children's Hospital

Robert D. MacDougall, M.Sc., DABR

Boston Children's Hospital

Harvard Medical School



Boston Children's Hospital
Until every child is well™



HARVARD MEDICAL SCHOOL
TEACHING HOSPITAL

Disclosures



Boston Children's Hospital
Until every child is well™



**HARVARD MEDICAL SCHOOL
TEACHING HOSPITAL 2**

Outline

1. Technical Implementation
 - Decisions: Data Format and Communication
 - Building a team and role of the physicist
2. Size-Specific and Indication-Based Diagnostic Reference
 - Regulatory
 - Protocol Optimization
3. Educating Users



Outline

1. Technical Implementation

- Decisions: Data Format and Communication
- Building a team and role of the physicist

2. Size-Specific and Indication-Based Diagnostic Reference

- Regulatory
- Protocol Optimization

3. Educating Users



Decisions – Data Format

1. RDSR: Event level data, allows for optimization
2. Secondary Capture/OCR
3. Image meta-data: OK for DX, NOT for CT, XA
4. MPPS and Custom Reports: The Last resort

Radiation Dose Structured Report

Patient:	
Referring Physician:	
Study: Pediatric/Pediatric Cardiac Catheterization	
Manufacturer: Siemens	
Completion Flag: PARTIAL	
Verification Flag: UNVERIFIED	
Content Date/Time:	

X-Ray Radiation Dose Report

Procedure reported

Projection X-Ray (113704, DCM)

HAS CONCEPT MOD

Has Intent:

Combined Diagnostic and Therapeutic Procedure (R-002E9, SRT)

Observer Type

Device (121007, DCM)

Device Observer UID

Device Observer Name

ARTIS123031

Device Observer Manufacturer

Siemens

Device Observer Model Name

AXIOM-Artis

Device Observer Serial Number

123031

Scope of Accumulation

Study (113014, DCM)

HAS PROPERTIES

Study Instance UID:

Accumulated X-Ray Dose Data

Acquisition Plane:

Plane A (113620, DCM)

Calibration:

Dose Measurement Device:

Dosimeter (A-2C090, SRT)

Calibration Date:

02-11-2015 21:05:23.000

Calibration Factor:

1

Calibration Uncertainty:

5 Percent

Calibration Responsible Party:

Siemens

Dose Area Product Total:

0.00152384 Gy.m2

Dose (RP) Total:

0.16296 Gy

Fluoro Dose Area Product Total:

0.00075835 Gy.m2

Fluoro Dose (RP) Total:

0.06333 Gy

Total Fluoro Time:

828 s

Acquisition Dose Area Product Total:

0.00076549 Gy.m2

Acquisition Dose (RP) Total:

0.09963 Gy

Total Acquisition Time:

75 s

Reference Point Definition:

15cm from Isocenter toward Source (113860, DCM)

Accumulated X-Ray Dose Data

Acquisition Plane:

Plane B (113621, DCM)

Calibration:

Dose Measurement Device:

Dosimeter (A-2C090, SRT)

Calibration Date:

02-11-2015 21:41:18.000

Calibration Factor:

1

Calibration Uncertainty:

5 Percent

Calibration Responsible Party:

Siemens

Dose Area Product Total:

0.00327012 Gy.m2

Dose (RP) Total:

0.36952 Gy

Fluoro Dose Area Product Total:

0.00189613 Gy.m2

Fluoro Dose (RP) Total:




Radiation Dose Structured Report

Type	Protocol	DAP [mGy-cm2]	Reference Point	Reference Point Dose [mGy]	Beam On Time [r]	kVp	mA	mAs	Start Time	Primary Angle [deg]	Secondary Angle [deg]	Fluoro Mode	Pulses per Secor	Number of Pulses
Fluoroscopy	FL <30kg 7.5	170.1	15cm from Isocenter	1.97	442	67	233.9	103.383	2016-07-26 14:56	91.2	-0.2	Pulsed	7.5	130
Fluoroscopy	FL <30kg 7.5	46.2	15cm from Isocenter	0.54	122.4	67	231.9	28.384	2016-07-26 14:56	91.2	-0.2	Pulsed	7.5	36
Fluoroscopy	FL <30kg 7.5	7.8	15cm from Isocenter	0.090000	22.8	67	208.8	4.76	2016-07-26 14:56	91.2	-0.2	Pulsed	7.5	6
Fluoroscopy	FL <30kg 7.5	417.9	15cm from Isocenter	4.83	1084.6	67	234.7	254.555	2016-07-26 14:56	91.2	-0.2	Pulsed	7.5	319
Fluoroscopy	FL <30kg 7.5	246.7	15cm from Isocenter	2.06	1084.6	64	176.2	191.106	2016-07-26 14:56	0.6	-0.3	Pulsed	7.5	319
Fluoroscopy	FL <30kg 7.5	113	15cm from Isocenter	1.31	272	69	233.3	63.457	2016-07-26 14:56	91.2	-0.2	Pulsed	7.5	80
Fluoroscopy	FL <30kg 7.5	635.3	15cm from Isocenter	7.59	966	69	237.8	229.714	2016-07-26 14:56	91.2	-0.2	Pulsed	7.5	210
Fluoroscopy	FL <30kg 7.5	382.5	15cm from Isocenter	4.62	927	69	237.1	219.791	2016-07-26 15:00	91.2	-0.2	Pulsed	7.5	206
Fluoroscopy	FL <30kg 7.5	28.3	15cm from Isocenter	0.34	68.4	70	228.5	15.629	2016-07-26 15:01	91.2	-0.2	Pulsed	7.5	18
Fluoroscopy	FL <30kg 7.5	83.3	15cm from Isocenter	1.01	205.2	69	234.1	48.037	2016-07-26 15:01	91.2	-0.2	Pulsed	7.5	57
Stationary Acquis	Card <30kg	314.9	15cm from Isocenter	3.8	346.8	70	319.5	110.802	2016-07-26 15:01	91.2	-0.2		30	102
Stationary Acquis	Card <30kg	170	15cm from Isocenter	1.44	346.8	66	323.2	112.085	2016-07-26 15:01	0.6	-0.3		30	102
Fluoroscopy	FL <30kg 7.5	376.9	15cm from Isocenter	4.55	927.5	69	235.1	218.055	2016-07-26 15:02	91.2	-0.2	Pulsed	7.5	265
Stationary Acquis	Card <30kg	944	15cm from Isocenter	11.41	914.6	70	367.4	336.024	2016-07-26 15:02	91.2	-0.2		30	269
Stationary Acquis	Card <30kg	472.2	15cm from Isocenter	4	911.2	66	341.6	311.265	2016-07-26 15:02	0.6	-0.3		30	269
Fluoroscopy	FL <30kg 7.5	10.1	15cm from Isocenter	0.090000	39.6	64	192.6	7.626	2016-07-26 15:03	0.6	-0.3	Pulsed	7.5	11
Fluoroscopy	FL <30kg 7.5	146	15cm from Isocenter	1.76	360.5	69	233.6	84.212	2016-07-26 15:03	91.2	-0.2	Pulsed	7.5	103
Fluoroscopy	FL <30kg 7.5	471	15cm from Isocenter	5.69	1172.5	69	234.9	275.42	2016-07-26 15:03	91.2	-0.2	Pulsed	7.5	335
Stationary Acquis	Card <30kg	3.9	15cm from Isocenter	0.050000	4.6	70	288.1	1.325	2016-07-26 15:04	91.2	-0.2		30	1
Stationary Acquis	Card <30kg	1.5	15cm from Isocenter	0.010000	4.2	66	297.9	1.251	2016-07-26 15:04	0.6	-0.3		30	1
Stationary Acquis	Card <30kg	428.3	15cm from Isocenter	5.17	387.6	70	395.9	153.45	2016-07-26 15:04	91.2	-0.2		30	116




Radiation Dose Structured Report


Accumulated RP Dose by Protocol




Dose Rate by Protocol



Peak Skin Dose Mapping



Toshiba (Real-time)



Radimetrics (Retrospective)



Boston Children's Hospital
Until every child is well™



**HARVARD MEDICAL SCHOOL
TEACHING HOSPITAL**

Secondary Capture: Requires OCR

Exam Protocol

Patient Info:

Name:

Sex: F ID:

Patient Position: HFS

Accumulated exposure data

Phys:

TotalFluoro: 5.8min

A Fluoro: 5.8min

B Fluoro: 0.0min

139.2pGym²

0.0pGym²

17.7mGy

0.0mGy

Exposures: 0

Total: 139.2pGym² 17.7mGy

Total: 139.2pGym² 17.7mGy

Total: 0.0pGym² 0.0mGy



DICOM Image Meta-Data



KVP	i	>	1514	0018,0060	DS	4 B	125
DeviceSerialNumber	i	>	1526	0018,1000	LO	14 B	72567767433727
SoftwareVersions	i	>	1548	0018,1020	LO	8 B	4.1.3.2
ProtocolName	i	>	1564	0018,1030	LO	6 B	Chest
SpatialResolution	i	>	1578	0018,1050	DS	6 B	0.148
DistanceSourceToDetector	i	>	1592	0018,1110	DS	4 B	1752
DistanceSourceToPatient	i	>	1604	0018,1111	DS	4 B	1698
EstimatedRadiographicMagnificationFactor	i	>	1616	0018,1114	DS	16 B	0.99376063528071
ExposureTime	i	>	1640	0018,1150	IS	2 B	2
Exposure	i	>	1650	0018,1152	IS	2 B	1
ExposureInuAs	i	>	1660	0018,1153	IS	4 B	1400
ImageAndFluoroscopyAreaDoseProduct	i	>	1672	0018,115e	DS	6 B	0.409
ImagerPixelSpacing	i	>	1686	0018,1164	DS	12 B	0.148\0.148
Grid	i	>	1706	0018,1166	CS	2 B	IN
AcquisitionDeviceProcessingDescription	i	>	1716	0018,1400	LO	64 B	UNIQUE: S:200 L:4.0 FC d:
RelativeXRayExposure	i	>	1788	0018,1405	IS	4 B	372
ExposureIndex	i	>	1800	0018,1411	DS	4 B	372
PositionerType	i	>	1812	0018,1508	CS	4 B	NONE
CollimatorShape	i	>	1824	0018,1700	CS	12 B	RECTANGULAR
CollimatorLeftVerticalEdge	i	>	1844	0018,1702	IS	2 B	0
CollimatorRightVerticalEdge	i	>	1854	0018,1704	IS	4 B	2321
CollimatorUpperHorizontalEdge	i	>	1866	0018,1706	IS	2 B	0
CollimatorLowerHorizontalEdge	i	>	1876	0018,1708	IS	4 B	2520
ViewPosition	i	>	1888	0018,5101	CS	2 B	PA
DetectorTemperature	i	>	1898	0018,7001	DS	4 B	29.2
DetectorType	i	>	1910	0018,7004	CS	12 B	SCINTILLATOR



DICOM Image Meta-Data



- Use with CR/DX
- Does not contain accumulated information or info on rejected studies



MPPS and Custom Solutions

- MPPS typically used for scheduling with appended messages
- Contain very limited Radiation Dose module (time, DAP) based on archived images
- Manual Entry is more reliable



MPPS and Custom Solutions

- Other solution: OCR from printed report to virtual printer



Boston Children's Hospital
Until every child is well™



**HARVARD MEDICAL SCHOOL
TEACHING HOSPITAL**

Decisions – Data Transfer

- Send from Modality
 - Fast, reduce burden on PACS
- Send from PACS
 - Maintains complete database (# studies, etc)
 - Can retrieve data through query or sync



Role

Responsibility



Boston Children's Hospital
Until every child is well™



**HARVARD MEDICAL SCHOOL
TEACHING HOSPITAL**

Role

PACS Administrator

Responsibility

Maintain server, coordinate upgrades,
configure PACS forwarding and sync



Boston Children's Hospital
Until every child is well™



**HARVARD MEDICAL SCHOOL
TEACHING HOSPITAL**

Role	Responsibility
PACS Administrator	Maintain server, coordinate upgrades, configure PACS forwarding and sync
Service Engineer	Configure modality to send dose report, define station name and AE title



Role	Responsibility
PACS Administrator	Maintain server, coordinate upgrades, configure PACS forwarding and sync
Service Engineer	Configure modality to send dose report, define station name and AE title
Radiologist Division Head	Review data, approve protocol changes



Role	Responsibility
PACS Administrator	Maintain server, coordinate upgrades, configure PACS forwarding and sync
Service Engineer	Configure modality to send dose report, define station name and AE title
Radiologist Division Head	Review data, approve protocol changes
Department Administrator	Dedicate time, resources, \$\$\$



Role	Responsibility
PACS Administrator	Maintain server, coordinate upgrades, configure PACS forwarding and sync
Service Engineer	Configure modality to send dose report, define station name and AE title
Radiologist Division Head	Review data, approve protocol changes
Department Administrator	Dedicate time, resources, \$\$\$
RSO	Ensure compliance with regulations



Role	Responsibility
PACS Administrator	Maintain server, coordinate upgrades, configure PACS forwarding and sync
Service Engineer	Configure modality to send dose report, define station name and AE title
Radiologist Division Head	Review data, approve protocol changes
Department Administrator	Dedicate time, resources, \$\$\$
RSO	Ensure compliance with regulations
Medical Physicist	All of the above



Actual Responsibility:

Configuration, interpretation, education,
developer(?)

Examination Dose Notifications		
Patient ID	Accession Number	Details
		<p>VOL: Reference Point Dose Value: 1131.1 [mGy] Threshold: 1000.0 [mGy] Dose Source: Scanner Output Criteria: gender Study: Pediatric Biopsy/Coronary Angio Modality: XA Model: Station: Date/Time: Institution: Action Level:</p>



Outline

1. Technical Implementation
 - Decisions: Data Format and Communication
 - Building a team and role of the physicist
2. Size-Specific and Indication-Based Diagnostic Reference
 - Regulatory
 - Protocol Optimization
3. Educating Users



2. Size-Specific DRR's - CT

Joint Commission: “The [critical access] hospital establishes or adopts diagnostic computed tomography (CT) imaging protocols based on current standards of practice, which address key criteria including **clinical indication**, contrast administration, age (to indicate whether the patient is pediatric or an adult), **patient size** and body habitus, and the **expected radiation dose index range**”



2. Size-Specific DRR's - CT

Joint Commission: “The [critical access] hospital reviews and analyzes incidents where the radiation dose index (CTDIvol, DLP, or size specific dose estimate [SSDE]) from diagnostic CT examinations **exceeded expected dose index ranges** identified in imaging protocols”



DRR MAX SETTINGS (Boston Children's Hospital)

Date: 07-05-2016

ANATOMIC REGION		PATIENT SIZE	CTDI_{vol} (mGy, 32)	CTDI_{vol} (mGy, 16)
HEAD	(Class 1)	PEDIATRIC	-	40
		ADULT	-	75
HEAD	(Class 2)	PEDIATRIC	-	20
		ADULT	-	40
HEAD	(Class 3)	PEDIATRIC	-	10
		ADULT	-	20
ABDOMEN (Class 1)		PEDIATRIC	10	15
		ADULT	25	-
ABDOMEN (Class 2)		PEDIATRIC	5	8
		ADULT	16	-
ABDOMEN (Class 3)		PEDIATRIC	3	4
		ADULT	8	-
CHEST (Class 1)		PEDIATRIC	7	11
		ADULT	18	-
CHEST (Class 2)		PEDIATRIC	4	6
		ADULT	9	-



Boston Children's Hospital
Until every child is well™



**HARVARD MEDICAL SCHOOL
TEACHING HOSPITAL**

DC I

DC II


DC III



Relative Dose = 100%



Relative Dose = 50%




Relative Dose = 25%





Boston Children's Hospital
Until every child is well™





HARVARD MEDICAL SCHOOL
TEACHING HOSPITAL






















Max mA too low




Boston Children's Hospital
Until every child is well™




**HARVARD MEDICAL SCHOOL
TEACHING HOSPITAL**


XR - Chest



XR - Chest



XR-Abdomen




Boston Children's Hospital
Until every child is well™



HARVARD MEDICAL SCHOOL
TEACHING HOSPITAL

XR-Abdomen



Boston Children's Hospital
Until every child is well™



**HARVARD MEDICAL SCHOOL
TEACHING HOSPITAL**

Outline

1. Technical Implementation
 - Decisions: Data Format and Communication
 - Building a team and role of the physicist
2. Size-Specific and Indication-Based Diagnostic Reference
 - Regulatory
 - Protocol Optimization
3. Educating Users



3. Educating Users - Technologists



- Work together to prepare “standing” dashboards
- Set up email reports (e.g. monthly)
- Develop QA loop defining when to escalate events and recommended action



3. Educating Users - Administrators



- Demonstrate compliance with regulatory/accreditation standards



- What to show if JC shows up...



3. Educating Users - Radiologists



- Potential dangers of dose history and the gambler's fallacy
- QI tool (MOC, etc)
- Justification for re-training of best practices



Thank You

Questions?



Boston Children's Hospital
Until every child is well™



**HARVARD MEDICAL SCHOOL
TEACHING HOSPITAL**