


Pitfalls and Remedies in PET/CT imaging for RT planning

Tinsu Pan, Ph.D.

*M.D. Anderson Cancer Center
The University of Texas*

Outlines


- Background
- Average CT (< 1 mSv) to reduce mis-alignment of PET and CT
 - Automatic processing without operator interaction
- Two different implementations
 - Average CT w/o gating (this talk)
 - Average CT w/ gating (GE/Siemens/Philips)
- Summary




PET/CT applications at M.D.Anderson

- Majority for diagnosis and staging of cancer
- One PET/CT scan per patient prior to treatment

Diagnostic Imaging




Radiation Oncology




*Free-breathing or breath-hold in CT
of PET/CT ?*

Helical CT data acquisition




Breathing Artifacts



Protocol: 16x0.625 mm, 0.8 s gantry rotation, pitch 1.375:1

Speed: 13.75 mm/0.8 s or 17.2 mm/s

Breathing artifacts to physiological info




$$\text{Breath cycle} = 80.3 / (13.75 / 0.8) = 4.67 \text{ s}$$

$$\text{Heart rate} = (21 / (13.75 / 0.8))^{-1} * 60 = 49 \text{ bpm}$$


Registration of Free-Breathing CT and PET maybe like




Misalignment in breathing states



Mis-matched PET-CT data sets



Mismatch 1:
CT diaphragm position
lower than PET





Mismatch 2:
CT diaphragm position
higher than PET

*Deep inspiration breath-
hold PET/CT*

Deep inspiration breath-hold (3 min)


FB PET/CT




DIBH PET/CT
5 sec DBH CT
9 x 20 sec DIBH PET

Deep inspiration breath-hold (20 s)


FB PET/CT




$\text{SUV}_{\max} = 3.6$



DIBH PET/CT



$\text{SUV}_{\max} = 8.1$



4D PET/CT

*(better with list-mode, and challenging to
perform in the clinic)*

Respiratory gating devices

Varian (Optical)




Anzai (Pressure)



Philips (air bellows)




Radiation Oncology




Diagnostic Imaging



4D PET patient study




SUV=5.0



SUV=8.5

Near end-expiration
PET/CT

Near end-expiration PET



94

COMPUTER METHODS AND PROGRAMS IN BIOMEDICINE 92 (2008) 90–98




Fig. 4 – Sagittal cross-sections of images from Patient 1 for (a) CT-std (left) and Ungated PET (right) images and (b) BH-CT (left) and BH-CT-based PET (right) images. Dotted lines are placed on the upper and lower boundaries of the lesion on the CT images and carried over onto the corresponding PET image. Cardiac motion artifacts are visible on both CT images because no cardiac gating was performed.


Average CT (ACT) < 1 mSv

*Data acquired at high temporal resolution and
averaged over one breath cycle*


Attenuation correction, RT dose calculation, IGRT

4D-CT

CT Scanner




Cine CT images



4D-CT




Monitoring system




Respiratory Trace

Sorting

CT Scanner




Cine CT images



4D-CT




Tracking system




Respiratory
Trace

Average CT (ACT) < 1 mSv

Sorting



MIP(mip)




Average

Tumor contouring with MIP CT



Soft tissue (400,40)



40% SUV_{max}



Lung (1000,-700)

20% SUV_{max}


Cine CT Data Acquisition




Average CT 5 mGy (0.87 mSv per 14 cm)

Average CT is better than slow CT

(2 adjacent CT slices of 2.5 mm apart)



Average CT



Average CT

Creating combined CT



ACT



Helical CT w ACT




Helical CT w/o thorax



Helical CT


Average CT for RT dose calculation




*Clinical examples of
average CT*

Clinical Studies


NSCLC




Esophageal






Lung tumor 1











Lung tumor 2




Lung tumor or lymph node




Lung tumor




Liver lesion



Kidney or liver uptake






Lung or liver lesion?




Lung lesion or liver lesion?


Average CT



Colorectal cancer




Colorectal cancer




Average CT

Inside the liver?



Outside the liver




Average CT

Tumor and cardiac imaging

HCT


PET_{HCT}



SUV=2.6






ACT

PET_{ACT}



SUV=5.0


Average CT in cardiac PET




HCT- misregistration

ACT

Improve the restaging after chemo




PET/CT scan indicated a positive response to induction chemo with HCT.



The patient had a negative response to the chemo with ACT.

Impact on treatment planning

Previous GTV was outlined based on CT and clinical PET without motion correction.
New GTV was redefined based on the correct information from PET with ACT.



Oncology PET/CT dose estimate

Injection dose: 10 mCi per patient

$$1 \text{ mCi} = 3.7 \times 10^7 \text{ dps} = 3.7 \times 10^7 \text{ bequerel (Bq)}$$

Radiation dose: 7 mSv from PET

5 to 10 mSv from CT


< 0.5 to 1 mSv from ACT

Radiation dose: 3.6 mSv from the environment


*Why are we not using
average CT?*

Complexity, Workflow and Cost

CT Scanner




Cine CT images



4D-CT





Tracking system




Respiratory
Trace

Average CT (ACT) < 1 mSv


MIP(mip)




Sorting



4D-CT Workflow



Cine CT Workflow without Gating





Operator Information

Exam Rx Image Works Manual



iLinq



PET InSite Service Shutdown

- Autovoice Volume
- Check Security
- Unix Shell
- Turn Off Extend HU
- Quick Snap

February 4 18:51

512: 114713 256: 458854



PET Partial Ready



Idle



Removed Series 999/812



Sent: 10328/3/... (dstaw1)



**** Film Composer started ****

Prescribed slice(s) are not visible over
localizer window.

Examinations :

Exam	Name	Date	Description	Mod1	Fmt	PPS	A	I
10017	zzdst2_12032001	Dec 03 08		CT	DIC01	-	I	N
10016	zzzgetest	Dec 02 08		CT	DIC01	-	I	N
10015	ZZZGETEST	Dec 02 08		CT	DIC01	-	I	N
10008	zzpetct	Dec 02 08		CT	DIC01	-	I	N
9991/DST2	ZZPETCT zzmont	Nov 24 08		CT	DIC01	-	I	N
9990/DST2	ZZPETCT zzmont	Nov 24 08		CT	DIC01	-	I	N
9979/DST2	zzdst2_112408	Nov 24 08		CT	DIC01	-	I	N
9973/DST2	zzdst2_112108	Nov 21 08		CT	DIC01	-	I	N
9965/DST2	zzdst2_112008	Nov 20 08		CT	DIC01	-	I	N
9957/DST2	zzdst2_111908	Nov 19 08		CT	DIC01	-	I	N
9947/DST2	zzdst2_111808	Nov 18 08		CT	DIC01	-	I	N
9939/DST2	ZZDST2_111708	Nov 17 08		CT	DIC01	-	I	N
9936/DST2	zzdst2_111408	Nov 14 08		CT	DIC01	-	I	N
9927/DST2	ZZDST2_111308	Nov 13 08		CT	DIC01	-	I	N
9920/DST2	zzdst2_111208	Nov 12 08		CT	DIC01	-	I	N
9919/DST2	zzzgetestdst2	Nov 11 08		CT	DIC01	-	I	N
999/DST2	Anonymous999	May 25 04	PET.NSCLC INI CT	DIC01	-	I	N	

164 examinations

one series

Series no 1 - PROSP

Img	Im Ctr	Thick	Gntry	Im Ctr	Im Ctr	SFOV	DFOV	Alg	Matrix	Midscan	Ph	Archive	
	I S-I (mm)	(mm)	(deg)	R-L	A-P	(cm)	(cm)		(sec)	XR			
1	I255.00	2.500	0.0	R	A	0.0	50.0	50.0	STND	512	0.250	I	No
2	I257.50	2.500	0.0	R	A	0.0	50.0	50.0	STND	512	0.250	I	No
3	I260.00	2.500	0.0	R	A	0.0	50.0	50.0	STND	512	0.250	I	No
4	I262.50	2.500	0.0	R	A	0.0	50.0	50.0	STND	512	0.250	I	No
5	I265.00	2.500	0.0	R	A	0.0	50.0	50.0	STND	512	0.250	I	No
6	I267.50	2.500	0.0	R	A	0.0	50.0	50.0	STND	512	0.250	I	No
7	I270.00	2.500	0.0	R	A	0.0	50.0	50.0	STND	512	0.250	I	No
8	I272.50	2.500	0.0	R	A	0.0	50.0	50.0	STND	512	0.250	I	No
9	I255.00	2.500	0.0	R	A	0.0	50.0	50.0	STND	512	0.401	I	No
10	I257.50	2.500	0.0	R	A	0.0	50.0	50.0	STND	512	0.401	I	No
11	I260.00	2.500	0.0	R	A	0.0	50.0	50.0	STND	512	0.401	I	No
12	I262.50	2.500	0.0	R	A	0.0	50.0	50.0	STND	512	0.401	I	No
13	I265.00	2.500	0.0	R	A	0.0	50.0	50.0	STND	512	0.401	I	No

992 images

- Autovoice Volume
- Check Security
- Unix Shell
- Turn Off Extend HU
- Quick Snap
- Add / Sub
- Edit Patient
- Film
- Mini Viewer
- Reformat
- Viewer



Exam Rx Image Works Manual



PET InSite Service Shutdown

February 4 18:50

512: 114621 256: 458486



PET Partial Ready



Idle



Removed Series 999/809



Sent: 10328/3/... (dstaw1)



**** Film Composer started ****

Prescribed slice(s) are not visible over localizer window.

Autovoice Volume
 Check Security
 Unix Shell
 Turn Off Extend HU
 Quick Snap

Exam	Name	Date	Description	ModI	Fmt	PPS	A	I
10017	zzdst2_12032001	Dec 03 08		I	CT	DICOI	-	I N
10016	zzzgetest	Dec 02 08		I	CT	DICOI	-	I N
10015	ZZZGETEST	Dec 02 08		I	CT	DICOI	-	I N
10008	zzpetct	Dec 02 08		I	CT	DICOI	-	I N
9991/DST2	ZZPETCT zzmont	Nov 24 08		I	CT	DICOI	-	I N
9990/DST2	ZZPETCT zzmont	Nov 24 08		I	CT	DICOI	-	I N
9979/DST2	zzdst2_112408	Nov 24 08		I	CT	DICOI	-	I N
9973/DST2	zzdst2_112108	Nov 21 08		I	CT	DICOI	-	I N
9965/DST2	zzdst2_112008	Nov 20 08		I	CT	DICOI	-	I N
9957/DST2	zzdst2_111908	Nov 19 08		I	CT	DICOI	-	I N
9947/DST2	zzdst2_111808	Nov 18 08		I	CT	DICOI	-	I N
9939/DST2	ZZDST2_111708	Nov 17 08		I	CT	DICOI	-	I N
9936/DST2	zzdst2_111408	Nov 14 08		I	CT	DICOI	-	I N
9927/DST2	ZZDST2_111308	Nov 13 08		I	CT	DICOI	-	I N
9920/DST2	zzdst2_111208	Nov 12 08		I	CT	DICOI	-	I N
9919/DST2	zzzgetestdst2	Nov 11 08		I	CT	DICOI	-	I N
999/DST2	Anonymous999	May 25 04	PET.NSCLC INI CT	I	DICOI	-	I	N

164 examinations

4 series

Series no 1 - PROSP

Img	Im Ctr	Thick	Gantry	Im Ctr	Im Ctr	SFOV	DFOV	Alg	Matrix	Midscan	Ph	Archive	
	I S-I (mm)	(mm)	(deg)	I R-L	I A-P	(cm)	(cm)			(sec)	%R		
1	I255.00	2.500	0.0	I R	I A	0.0	50.0	I STND	512	0.250	I	No	
2	I257.50	2.500	0.0	I R	I A	0.0	50.0	I STND	512	0.250	I	No	
3	I260.00	2.500	0.0	I R	I A	0.0	50.0	I STND	512	0.250	I	No	
4	I262.50	2.500	0.0	I R	I A	0.0	50.0	I STND	512	0.250	I	No	
5	I265.00	2.500	0.0	I R	I A	0.0	50.0	I STND	512	0.250	I	No	
6	I267.50	2.500	0.0	I R	I A	0.0	50.0	I STND	512	0.250	I	No	
7	I270.00	2.500	0.0	I R	I A	0.0	50.0	I STND	512	0.250	I	No	
8	I272.50	2.500	0.0	I R	I A	0.0	50.0	I STND	512	0.250	I	No	
9	I255.00	2.500	0.0	I R	I A	0.0	50.0	I STND	512	0.401	I	No	
10	I257.50	2.500	0.0	I R	I A	0.0	50.0	I STND	512	0.401	I	No	
11	I260.00	2.500	0.0	I R	I A	0.0	50.0	I STND	512	0.401	I	No	
12	I262.50	2.500	0.0	I R	I A	0.0	50.0	I STND	512	0.401	I	No	
13	I265.00	2.500	0.0	I R	I A	0.0	50.0	I STND	512	0.401	I	No	

992 images

Exam no 999, May 25 04, Anonymous999

Ser	Type	Imgs	Description	ModI	PPS	I
1	PROSP	992	I	I CT	I	-
810	PROSP	32	I P_AVG	I CT	I	-
811	PROSP	32	I P_mip	I CT	I	-
812	PROSP	32	I P_MIP	I CT	I	-

Automatically generated

Add / Sub
 Edit Patient
 Film
 Mini Viewer
 Reformat
 Viewer

Summary

- Respiratory gating PET can improve quantification, but is challenging to be performed in the clinic
- Average CT can improve registration of CT and PET, is used for dose calculation and IGRT in radiation therapy
- MIP CT can assist tumor contouring
- Workflow and efficiency can be improved, and cost can be reduced