

# Frameless SRS Using CyberKnife

*Jun Yang, PhD*  
*Jing Feng, MS*

CyberKnife Center of Philadelphia  
June, 2019  
AAPM Annual Meeting



# Study Objectives

- Components and Work Flow of CyberKnife
- Motion Management of SRS Cases
- Dosimetry Characteristics
- New Development of CyberKnife
- QA



## ✓ CyberKnife Components


Motion management

Dosimetry

Quality Assurance






# CyberKnife<sup>®</sup> Components




# Linear Accelerator

- 330 lbs.
- 6 MV X-ray
- 1000 MU/min  $\pm$  10% at 800 mm SAD
- Three set collimators
  - 5 – 60 mm circular collimators
  - 5 – 60 mm dodecagonal (12-sided) IRIS variable aperture collimators
  - 100 x 115 mm Incise™ 2 MLC





# Robot Specifications





- **6 axis / joint motion**
- **1,220 Kg**
- **300 Kg payload**
- **208 VAC, 3 Phase (PDU)**
- **0-45 °C Operating Range**
- **<75% Relative Humidity**
- **41 m<sup>3</sup> Working Envelope**
- **2500 mm Maximum Reach**
- **0.12 mm Repeatability**

Made by KUKA of Germany



# Robot Motion During Treatment

- Nodes assigned
  - Approximately 130
  - 12 directions each node
- ~1500 beams total (Simplex)
- Up to 5000 beams (Sequential)
- Treatment path



# Treatment Couch

## Axum Couch



- Accommodates up to 159 Kg patient (350 lbs)
- Motorized control with 5 DOF (no YAW)
- Repeatability: 0.3mm/0.3°

## RoboCouch®





- Accommodates up to 227 Kg patient (500 lbs)
- Motorized control with 6 DOF
- Repeatability: 0.1mm/0.1°





# Image Tracking System

- Two Diagnostic X-Ray sources
- Two ASi image detectors
- Patient imaged at 45° orthogonal angles
- Image Center: 920mm  $\pm$  5mm from floor
- Real-time, live images compared against DRRs generated from CT
- During treatment, Robot adjusts position based on the comparison



# Development of CyberKnife


## G4 Analog




## M6 Digital




Fixed Collimator




IRIS™ Collimator



Incise™ 2 MLC



# Incise™2 MLC



- Number of Leaves: 52 leaf pair
- Leaf Width(at 800 mm SAD): 3.85 mm
- Max Treatment Field Size(at 800 mm SAD): :115 mm x 100 mm
- Leaf Position Accuracy(at 800 mm SAD):  $< \pm 0.95$  mm
- Max Leaf Speed(at 800 mm SAD): 25mm/s
- Leaf Height: 90.0 mm
- Leaf Tip Design: Three Flat, Focused Edges
- Leaf Side Design: Flat, Focused
- Source to Collimator Distance: 400.0 mm
- Transmission:  $< 0.3\%$  average ( $< 0.5\%$  Max) relative to 100mm x 100mm field at 800mm
- Penumbra:  $< 3.5$  mm for 10mm x 10 mm  
 $< 12$  mm(X)& $20$  mm(Y) for 100mm x100 mm
- MLC Weight: 54 Kg



## CyberKnife Components

✓ Motion management

Dosimetry

Quality Assurance





# Tracking Methodology

- DRR (Digitally Reconstructed Radiographs from planning CT) pairs used as references
- X-ray image pairs acquired in real time
- Registration between DRR and X-ray images
- The patient's rigid transformation calculated





# Case Specific Tracking Modalities

- ✓ Skull Tracking ----- Intracranial tumor
- ❑ XSight (spine) Tracking ----- Spine tumor
- ❑ Fiducial Tracking ----- Soft tissue
- ❑ Synchrony Tracking ----- Moving Soft tissue
- ❑ XSight Lung ----- Moving visible lung tumor
- ❑ Lung Optimization Treatment ----- a full set of tracking for lung tumor without fiducial



# Skull Tracking and Correction




Image Tracking & Robotic Correction

# Targeting Accuracy

- Mechanical accuracy
  - 0.12 mm (Kuka Specification 2004)<sup>1</sup>
- Targeting accuracy for targets not affected by respiration
  - 0.95 mm (Xsight® Specification)
  - 0.52 +/- 0.22 mm (Muacevic et. Al. 2006)<sup>2</sup>
  - 0.49 +/- 0.22 mm (Ho et. al. 2008)<sup>3</sup>
  - 0.4 +/- 0.2 mm (Antypas and Pantelis 2008)<sup>4</sup>
  - 0.47 +/- 0.24 mm (Drexler & Furweger 2009)<sup>5</sup>
- Targeting accuracy for targets that move with respiration
  - 1.5 mm (Synchrony® Respiratory Tracking System specification)
  - 0.70 +/- 0.33 mm (Dieterich et. Al. 2004)<sup>6</sup>
  - 0.47 +/- 0.24 mm (Drexler and Furweger 2009)<sup>5</sup>



CyberKnife Components

Motion management

Dosimetry

Quality Assurance



# Immobilization & Simulation



Thermoplastic Mask, Supine

CT without contrast, ~ 1 mm slice thickness, contiguous scan from 2 cm above the superior end of skull to 2 cm inferior to the chin.



T1 weighted MRI scan with contrast, 1 - 2 mm slice thickness, covering the lesion(s), orbits and all ventricles



# Typical Prescription Doses (Ref Only)



- Brian Mets (Per RTOG 0320)
  - <2.0cm 24 Gy x 1
  - 2.1-3.0 cm 18 Gy x 1
  - 3.1-4.0 cm 15 Gy x 1
- Surgical Bed: 7~9Gyx3, 4~6Gyx5
- Acoustic Neuroma : 6~8Gy x3, 12~14Gy x1
- Trigeminal Neuralgia: 60~90GyX1 Max  
(Cone, Iris, MLC?)
- Meningioma: 6~8Gyx3, 12~18Gy x 1
- Typical PTV margin  $\leq 1$ mm except Surgical bed



# Dosimetry: Various Beam Arrangements


- **Isocentric plan**

- Fast dose fall-off
- Limited conformality




- **Non-isocentric**

- Excellent conformality
- Dose fall-off not as fast




# Isocentric Beam Weight Adjustment



**Sphere Shape Dose  
Equal MU per Beam**



**Height > Width  
Oval Shape Dose**




**Height < Width  
Oval Shape Dose**

**Isoconformal  
Unequal MU per Beam  
Limitedly Improved conformiaty**






# Isocentric Case: Brian mets






# Multi-Isocentric Case: Brian mets






# Non-Isocentric Case: Brian mets





# Isocentric + Non-Isocentric Case





# Dose Heterogeneity VS Dose Fall off in Isocentric plan




70% ≈ 60% > 50% > 80% >> 90%

Absolute Dose profiles, with 20 Gy Rx at Different isodose Lines

Note: Max Dose = 100%

# Dose Heterogeneity VS Dose Fall off in Non-Isocentric plan



Note: Max Dose = 100%

CyberKnife Components

Motion management

Dosimetry

Quality Assurance





# Absolute Output: TG 51

- 1MU = 1 cGy for 60 mm fixed cone at 800mm SAD with 15 mm Depth;
- $\%dd(10)_x$  measured at 1000 mm SSD with 100 mm Depth;
- Calculated Equivalent Square;
- Interpolate to 800 mm SAD using BJR data;



# QA guidelines: **TG 135** Quality assurance for robotic radiosurgery. 2011 (New Version coming)

## IV.B. Daily QA

Section	Item	Tolerance
II.A.2	Safety interlocks (Door, console EMO, Key)	Functional
	CCTV cameras and monitors	Functional
	Audio monitor	Functional
	Collimator assembly collision detector	Functional
II.B.1	Accelerator warm-up: 6000 MU for open chambers, 3000 MU for sealed chambers	N/A
	Accelerator output	<2%: no change needed >2%: adjust calibration
	Detection of incorrect and missing secondary collimator	Functional
III.B.2	Visual check of beam laser and a standard floor mark.	<1 mm
III.C.1	AQA test	< 1 mm from baseline

## IV.C. Monthly QA

Section	Item	Tolerance
II.A.2	Safety interlocks.	Functional
II.B.2	Energy constancy.	2%
	Beam symmetry.	>3%
	Beam shape.	>2% Compared to beam data
	Output.	> 2%
II.C.1	Imager alignment.	1 mm or center pixel $\pm$ 2 pixels
II.C.3	Contrast, noise, and spatial resolution of amorphous silicon detector.	To be decided by user based on available literature
	Homogeneity/bad pixels.	
II.D	Custom CT model: CT QA (spatial accuracy, electron density).	See TG 66 (Ref. 29)
III.B.1	Verify relative location of beam laser vs. radiation CAX has not changed.	0.5 mm
III.B.2	Visually check isocentric plan to verify beam laser illuminates isocrystal; rotate through path sets each month	Laser on isocrystal for each node
III.C.2	Intracranial and extracranial E2E; set schedule to cycle through each clinically used tracking method and path.	<0.95 mm or <1.5 mm for motion tracking
III.C.3	Nonisocentric patient QA or DQA; ideally performed quarterly.	DTA 2 mm/2%; Synchrony DTA 3%/3 mm
III.D	Observe Synchrony treatment or simulation; listen for unusual noise and visually check for vibrations.	No significant change

## IV.D. Annual QA

Section	Item	Tolerance
II.A.2	EPO button	Functional
II.B.3	TG 51 or IAEA TRS-398, including secondary independent check.	Adjust calibration if >1% difference is found
	Beam data checks on at least three collimators, including largest and smallest collimator (TPR or PDD, OCR, output factors).	To be decided by user
	Dose output linearity to lowest MU/beam used.	1%
II.C.2	Imager kVp accuracy, mA station exposure linearity, exposure reproducibility, focal spot size.	See Table 1 for references
II.C.3	Signal to noise ratio, contrast-to-noise ratio, relative modulation transfer function, imager sensitivity stability, bad pixel count and pattern, uniformity corrected images, detector centering, and imager gain statistics.	Compare to baseline
II.D	TG 53 as applicable.	TG 53 (Ref. 26)
	CT QA (in addition to monthly).	See TG 66 (Ref. 29)
	Data security and verification.	Functional
III.B.2	2nd Order Path Calibration; currently only possible with the help of a service engineer.	Each node < 0.5 mm RMS < 0.3 mm
III.D	Check noise level of optical markers.	<0.2 mm
IV.C	Run Synchrony E2E test with at least 20 deg phase shift; analyze penumbra spread.	To be decided by user
IV.C	Monthly QA.	In addition to tolerances listed above, update all parameters and checklists
IV.B	Daily QA.	Update parameters



# Minimum QA guidelines: **TG 277** AAPM-RSS Medical Physics Practice Guideline 9.a. for SRS – SBRT

**Table 2: Minimum equipment QA and tolerances for robotic linac systems**

Frequency	Test	Tolerance
<b>Daily*</b> <i>*On days of clinical use</i>	Head laser alignment check	1.0 mm
	Safety interlocks	Functional
	Automatic QA (AQA) test* <i>*If the clinic has both fixed cones and Iris™ collimator, the AQA test should alternate between fixed cones and Iris™, with each system tested at least weekly</i>	Total targeting ≤ 1.0 mm from baseline, not exceeding manufacturer's specification
	Accelerator output constancy	± 3%
<b>Monthly</b>	Energy constancy	± 2%
	Beam symmetry, relative	± 3% for 40 mm field, ± 4% for 60 mm field,
	Accelerator output constancy	± 2%
	Imager alignment	1mm or center pixels ± 2 pixels
	Iris Field size spot check	0.5 mm, 3 or more field sizes ≥ 10 mm
	Picket fence for MLC ( <i>if applicable</i> )	Visual check
<b>Quarterly</b>	E2E localization assessment (Each tracking mode used clinically)	1.0 mm static target, 1.5 mm motion tracking
<b>Annually</b>	Emergency Power Off (EPO) button, safety interlocks	Functional
	Accelerator output	± 2.0%
	MU linearity (>10 MU to highest MU used clinically)	± 2%
	Path verification	≤0.5 mm maximum per node, ≤0.3 mm average
	Imager kVp accuracy, mA station exposure linearity, isopost alignment with center pixel	± 10%, ± 20%, and 1 mm respectively
	Beam laser and radiation beam alignment for cone, Iris and MLC	0.5 mm from baseline
	AQA baseline	Re-check AQA baseline
	Beam data verification - Relative output factors for cones, Iris and/or MLC covering the range used clinically	± 2% from baseline for > 1.0 cm apertures, ± 5% from baseline for ≤ 1.0 cm apertures



# Quality Assurance

- End 2 End test
- BB test
- AQA
- Beam analysis
- Plan Dose Verification
- Image system test
- Daily, Monthly, Quarterly and Annually QA



# Morning QA - Output





# Morning QA – AQA(W-L test)



First select the beam angle setting (default: 0 degrees). Choose the image source from the drop-down menu. Click Browse to open up the image file. Rotate and/or flip the image using the appropriate icons that the image position within the field. When ready, click the process button to find the displacements in the X, Y, Z coordinates of the imaging plane.

Select Image:

Image A  
Inf

Left Right

Sup

Image B  
Inf

Post Ant

Sub

Notes:

Date:  
09-Jan-2017

Time:  
10:46:45

**Image A Coordinates**

X centroid offset : -0.22611 mm  
Y centroid offset : 0.12663 mm

Eccentricity (Beam) : 0.14718  
Eccentricity (Shadow) : 0.20844

**Image B Coordinates**

X centroid offset : 0.012985 mm  
Y centroid offset : 0.38456 mm

Eccentricity (Beam) : 0.12375  
Eccentricity (Shadow) : 0.197

Average Center Pixel Intensity : 26678  
(Coarse Estimate)

**Patient Plane Coordinates**


X offset : -0.2556 mm  
(superior-inferior)

Y offset : 0.22611 mm  
(right-left)

Z offset : 0.012985 mm  
(posterior-anterior)

Radial Error : 0.3415 mm



ACCURAY




# Monthly QA- Output/Energy



# Monthly QA – Symmetry and Flatness/ Laser Alignment







Open the Film Image...

Pickup Background 36957.6544

Pickup Peak 20805.9909

Film Scan Resolution: 300 DPI

Contour Line: 50 %

Draw the Contours

Pick Laser Point

Result  
Center: Radium:  
X: 31.92 mm 21.29 mm  
Y: 27.73 mm

Laser from the Center  
Laser Point  
X: 0.00 mm  
Y: -0.17 mm






# Monthly QA – IRIS Collimator Field Sizes



The screenshot displays the Accuray Iris QA v1.0 software interface, which is used for quality assurance of iris collimators. The interface is divided into several sections:

- User Inputs:** Fields for Iris S/N, Scanner S/N (Unknown), Scanner dpi (300), Film Batch # (08261502), Blank Pixel Value (41774), OD Threshold (0.612), Film Orientation (Landscape), Notes, and Physicist (Jun Yang, Phd DABR).
- Scan Processing Pad:** Contains two image windows: "Imported Image" (a green-tinted scan of the collimator) and "Cropped Image" (a grayscale version of the same scan).
- Analysis Results:** Displays the following data:
  - Date: 09-Jan-2017
  - Iris S/N: Unknown
  - Scanner S/N: Unknown
  - Scanner dpi: 300
  - Film Batch #: 08261502
  - Blank Pixel Value: 41774.0
  - OD Threshold: 0.612
  - Scan orientation: Landscape
  - Image File: EZE\_20MM\_IRIS\_111716907.tif
  - Modification Date: 17-Nov-2016
  - Collimator Type: Iris
  - Nominal Size (mm): 20.0
  - Equivalent Diameter (mm): 20.07
  - Profile Diameter (mm): 20.07
  - Std. Dev. of Profile Diameter (mm): 0.15
  - Central PV: 21231.0
  - Notes: Enter Notes
  - Physicist: Jun Yang, Phd DABR
- Thresholded Image:** A black circle on a white background representing the thresholded OD.
- OD Contours:** A graph showing contours at 30%, 50%, 70%, and 90% levels. The x and y axes are labeled "X position (mm)" and "Y position (mm)" respectively, with values from 0 to 50. A central crosshair is present, and the text "20.07 mm\*" is displayed on the graph.
- Export Results:** A dropdown menu to "Select export format" and a "Save..." button.
- Buttons:** "Auto Threshold..." and "Run Analysis" buttons are located at the bottom of the Scan Processing Pad.



# Monthly QA – E2E





**End-to-End (E2E) Film Analysis**





**Image 1 Threshold Area Information**

Centroid Area:	140171
Pixels to Left:	376.13
Pixels to Top:	373.2
Eccentricity:	1.12



**Image 2 Threshold Area Information**

Centroid Area:	111116
Pixels to Left:	381.65
Pixels to Top:	375.96
Eccentricity:	1.36



E2E Reference		Calibration Information	
Test:	IRIS FUD	Hor pix/mm:	11.9
Date:	11/17/2016	Vert pix/mm:	11.81
Film Label:		Sup. Pos mm:	31.75
Scanner:	Epson 1680	Ant. Pos mm:	31.75
Film Type:	EBT 2	Left Pos mm:	31.75
Film Batch:			
Analysis Color:	RED		
Image Bit Depth:	48		

<b>Pixel Measurement</b>	Contour Threshold: 25272
Minimum: 19644	BG Film: 45489

**Image 1 (A/L Image) Info**

mm from left edge:	31.61
mm from anterior edge:	31.6
contour area/ball area:	1.26

**Error Information**

left error mm:	0.14
anterior error mm (A/L image):	0.15
superior error mm:	-0.32
anterior error mm (A/S image):	-0.08
average anterior error mm:	0.03
<b>TOTAL TARGETING ERROR mm:</b>	<b>0.35</b>

**Image 2 (A/S Image) Info**

mm from superior edge:	32.07
mm from anterior edge:	31.83
contour area/ball area:	1



# Monthly QA – Imaging Center





Thanks

