

**Scripting and Automation for
Efficient and Effective
Chart Checks in a
Pinnacle/Mosaik Environment**

Badal Juneja
Leonard Kim
Michael Walsh

MD Anderson
Cancer Center
Cooper
Making Cancer History


1

**Scripting and Automation for
Efficient and Effective
Chart Checks in a
Pinnacle*/Mosaik Environment**

* Pinnacle/RayStation

MD Anderson
Cancer Center
Cooper
Making Cancer History

2




3


- ✓ **What?**
- ✓ **Why?**
- ✗ **How?**

→ juneja-badal@cooperhealth.edu

4




5




Physicist sees the plan for the first time

6



7



8

What do we check?

- Check orders and simulation note and verify against treatment planning images (e.g. immobilization, compression, breath hold, PET/CT)
- Verify image orientation (supine or prone only)
- Check volumes (e.g. expansions/crops, target dose labeling vs. prescription)
- Verify necessity and accuracy of density/material overrides
- Verify patient support structure location
- Verify patient support structure accuracy
- Check DVH scorecard against prescription note
- Verify correct CT-density table used
- Verify treated area is immobilized
- Check CT field of view and extent (e.g. no patient cut-off)
- Verify correct outside-patient air threshold
- Verify localization point matches tattooed isocenter
- Verify accurate application of bolus
- Check appropriate location of isocenter relative to target
- If multi-target, assess number of isocenters to be used
- Verify same isocenter used for all beams
- Verify same machine used for all beams
- Evaluate possibility of collision
- Verify correct isocenter and dose calculation point
- Check for appropriate beam energy
- Verify 6 MV used for VMAT beams
- Check naming convention (including laterality, orientation)
- Verify partial arcs are used on correct side
- Verify arc directions are correct
- Verify sagittal arcs move away from patient's head
- Verify appropriate collimator angles for VMAT beams
- Check dose grid coverage
- Verify correct dose grid resolution
- Verify dose engine
- Verify homogeneity correction

9


**"Implementation and Evaluation of a TG-275 Key Recommendation:
An Additional Early Physics Check" – Table 1**

[illegible]

Source: <https://doi.org/10.1016/j.pmo.2022.05.011>

10

Are we **adding** or **replacing** a check?



11


FMEA: RPN = O · S · D

0 – Occurrence

S – Severity

D – Detectability (lack of)

12



13

Pre-check Helper – Pinnacle


Condition	Desired Value	Current Plan Value	Result
CT Density Table	Physics By Bone, PencilRayline	PhysicsRayline	Pass
Outside Patient Air Threshold	0.0	0.0	Pass
Localization Check	Plan Name: Carotid City WB v1.0.0 - 0.04 - 0.05 - 0.05	Local Coordinates: 0.04 - 0.05 - 0.05	Pass
Crash Density Overlay Order	Crash Density Overlay - 0.05	Crash Density Overlay - 0.05	Pass
Same ID	Same Instance for all Beams	MD Instance	Pass
Same Machine	Same Machine for all Beams	ONE	Pass
Same Prescription	Same Prescription for all Beams	0.000000	Pass
Dose Grid Resolution	0.015000	0.015000	Pass
Dose Calculation Engine	CC Calculation or Electron SC	CC Calculation	Pass
Heterogeneity Correction	All Beams Heterogeneous	Heterogeneous	Pass
Dynamic Arc: BeamType	Dynamic Arcs	Dynamic Arcs	Pass
Dynamic Arc: Energy	6.000 MeV	6.000	Pass
Dynamic Arc: Collimator	600 x 600 mm	600 x 600 mm	Pass
Dynamic Arc: Arc Length	0.0 to 360.0 degrees	360.0	Pass

14


Pre-check Helper – RayStation

Pre-Check Helper Results	
CT Series Description: Extended FOV	All beams use 6 MV
CT Density table is correct: Youzhou Philips Big Bone	Acceptable number of beams Beam: 2
Medium 'Water' couch fixed	Collimator angles not equal to 0
Couch material verified	Collimator angles differ
External contour exists: Eftbody	All arcs are coplanar
Machine is consistent with couch	Beam direction match normal direction
Localization point exists: MD Instance	Minimum arc length acceptable
All beam share common isocenter: (5.44, 2.06, 10.0)	Arc direction allowed
Dose grid resolution correct: 3 mm	No illegal arcs present
Beam direction: 0.000000, 0.000000, 1.000000	
Beam direction: 0.000000, 0.000000, 1.000000	

15




16



17

While automation holds great promise for improving the efficiency and effectiveness of plan and chart review, it is important to be aware of its limitations. If it is not properly implemented or tested it could lead to errors being **systematically unidentified**.

— TG 275



19

1. Data File			
Test	Expected	Current	Status
Energy matches between prostate and rectum	5 MV	5 MV	
MRs match between prostate and rectum	200	200.000	
IGRs match between prostate and rectum	81.5	81.5	
Isocenter match between prostate and rectum	< 0.5, 4.02, 40.10	< 0.5, 4.02, 40.10	
All gantry angles are same between prostate and rectum	GA Start = 180.0, Stop = 170	GA Start = 180.0, Stop = 170.0	
All collimator angles are same between prostate and rectum	CA 10	CA 10.0	
All couch angles are same between prostate and rectum	Couch Angle 0	Couch Angle 0.0	
All CP weights same between prostate and rectum	Comparing 10 CPs	Max Diff: 0.00000	

20

“What gets measured gets managed.”

— Peter Drucker

21

What do you check?

- Check orders and simulation note and verify against treatment planning images (e.g. immobilization, compression, breath hold, PET/CT)
- Verify image orientation (supine or prone only)
- Check volumes (e.g. expansions/crops, target dose labeling vs. prescription)
- Verify necessity and accuracy of density/material overrides
- Verify patient support structure location
- Verify patient support structure accuracy
- **Check DVH scorecard against prescription note**
- Verify correct CT-density table used
- Verify treated area is immobilized
- Check CT field of view and extent (e.g. no patient cut-off)
- Verify correct outside-patient air threshold
- Verify localization point matches tattooed isocenter
- Verify accurate application of bolus
- Check appropriate location of isocenter relative to target
- If multi-target, assess number of isocenters to be used
- Verify same isocenter used for all beams
- Verify same machine used for all beams
- Evaluate possibility of collision
- Verify correct isocenter and dose calculation point
- Check for appropriate beam energy
- Verify 6 MV used for VMAT beams
- Check naming convention (including laterality, orientation)
- Verify partial arcs are used on correct side
- Verify arc directions are correct
- Verify sagittal arcs move away from patient's head
- Verify appropriate collimator angles for VMAT beams
- Check dose grid coverage
- Verify correct dose grid resolution
- Verify dose engine
- Verify homogeneity correction

22

Scorecard

File


Name: **PTV_LPV_01_01_01** (SEL: Appr: PTV)

Description:

ROI	Type	Primary Goal Dose cGy	Primary Goal Volume	Units	Secondary Goal Dose cGy	Secondary Goal Volume	Units	Dose cGy	Volume at Primary Goal Dose	Units	Result
CTV66	Min DVH (%)	6600	98.600	%	0	0.000	%	6647.2	Min 100.000	%	Pass
PTV66	Min DVH (%)	6600	94.600	%	0	0.000	%	6517.5	Min 99.764	%	Pass
CTV67	Min DVH (%)	6700	86.600	%	0	0.000	%	6725.1	Min 99.861	%	Pass
PTV67	Min DVH (%)	6700	84.600	%	0	0.000	%	6764.4	Min 97.764	%	Pass
CTV64	Min DVH (%)	6400	98.600	%	0	0.000	%	6442.8	Min 100.000	%	Pass
PTV64	Min DVH (%)	6400	94.600	%	0	0.000	%	6305.2	Min 99.894	%	Pass
Brain	Max Dose	5000	0.000	%	0	0.000	%	357.0	Max ---	%	Pass
Brainstem	Max Dose	5400	0.000	%	0	0.000	%	699.2	Max ---	%	Pass
SpinalCord	Max Dose	4800	0.000	%	0	0.000	%	2442.8	Max ---	%	Pass
SpinalCord (HVS)	Max Dose	5000	0.000	%	0	0.000	%	3473.7	Max ---	%	Pass
ParotidL	Mean Dose	2500	0.000	%	0	0.000	%	2086.4	Mean ---	%	Pass
ParotidR	Mean Dose	2500	0.000	%	0	0.000	%	6121.5	Max 38.344	%	Pass
ParotidL	Mean Dose	2600	0.000	%	0	0.000	%	2232.0	Mean ---	%	Pass
ParotidR	Mean Dose	2600	0.000	%	0	0.000	%	6101.0	Max 38.922	%	Pass
Larynx	Mean Dose	3500	0.000	%	0	0.000	%	3380.4	Mean ---	%	Pass
Mandible	Max Dose	4200	0.000	%	7200	0.000	%	6820.7	Max ---	%	Pass

23

Auto-naming contours based on TG-263 guidelines



24



28



29



30
