




UPDATING TG-51

When will it end?

Malcolm McEwen
Ionizing Radiation Standards
National Research Council, Canada

AAPM Annual Meeting, Charlotte, 2012


What are these updates?

Working Group review recommends:

- **Photons:**
 - i. An updated list of chambers
 - ii. Review of calculated k_Q factors
 - iii. Uncertainty analysis
 - iv. Implementation guidance notes (clarification)
- **Electrons:** More widespread revision required

AAPM's TG-51 protocol for clinical reference dosimetry of high-energy photons and electron beams


James L. Scott, University of Toronto, Toronto, Ontario, Canada
 Peter J. Biggs, University of Toronto, Toronto, Ontario, Canada
 Richard Anderson, National Institute of Standards and Technology, Gaithersburg, Maryland, USA
 B. M. Coursey, National Institute of Standards and Technology, Gaithersburg, Maryland, USA
 M. F. Palmer, National Institute of Standards and Technology, Gaithersburg, Maryland, USA
 M. J. Stachurski, National Institute of Standards and Technology, Gaithersburg, Maryland, USA
 B. M. Coursey, National Institute of Standards and Technology, Gaithersburg, Maryland, USA
 M. J. Stachurski, National Institute of Standards and Technology, Gaithersburg, Maryland, USA
 B. M. Coursey, National Institute of Standards and Technology, Gaithersburg, Maryland, USA
 M. J. Stachurski, National Institute of Standards and Technology, Gaithersburg, Maryland, USA
 B. M. Coursey, National Institute of Standards and Technology, Gaithersburg, Maryland, USA
 M. J. Stachurski, National Institute of Standards and Technology, Gaithersburg, Maryland, USA



Part 1 - photon addendum

The report will cover the following:


- A. k_Q factors for new chambers
- B. Recommendations for implementation
- C. Uncertainty analysis for implementation of TG-51
- D. Comparison of measured and calculated k_Q factors



Part 1 - photon addendum


The report will cover the following:

- A. k_Q factors for new chambers
- B. Recommendations for implementation**
- C. Uncertainty analysis for implementation of TG-51
- D. Comparison of measured and calculated k_Q factors




TG-51 photons – what stays?

- Procedure remains unchanged
 - ✓ Continue to follow the procedure in the TG-51 document
- TG-51 remains based on a calibration coefficient obtained in Co-60
 - ✓ MV standards and calibration services are already available in certain countries but widespread dissemination in the US is not realistic at the present time.
- Calculated k_Q factors
 - ✓ Measured k_Q data are available for some chamber types
 - ✓ MV calibration services unlikely to demand in North America
- $\%dd(10)_x$ remains the beam quality specifier
 - ✓ See discussion later





B. Recommendations

1. Implementation of TG-51 Addendum
2. k_Q data sets
3. Reference-class ionization chamber
4. Choice of polarizing voltage
5. Measurement of polarity correction, P_{pol}
6. Effective point of measurement
7. Use of small volume chambers in relative dosimetry
8. Non-water phantoms
9. Application to flattening-filter-free linacs



B.1 Implementation of Addendum

- **The addendum should be implemented!**
- Minor changes in experimental procedure
- New equipment may be required
- Development of uncertainty budget may take some time






B.2 k_Q data sets

1. For chambers listed in both this addendum and the original TG-51 protocol, the k_Q factors listed in the addendum should be used.
2. For chambers that are not listed in either the original TG-51 protocol or in this addendum the recommendations of Section XI of TG-51 should be followed.

XI. USING OTHER ION CHAMBERS

This protocol provides k_Q data for the vast majority of chambers used in clinical reference dosimetry in North America as evidenced by the data on ADCL calibrations. However, other cylindrical chambers can be used by finding the closest matching chamber for which data are given. The critical features are, in order, the wall material, the radius of the air cavity, the presence of an aluminum electrode, and the wall thickness. As long as the wall material is matched and the chamber is "normal," these matching data should be accurate to within 0.5%. It is the responsibility of the user to confirm this by comparing the results to those of a calibrated cylindrical chamber for which data are given in the protocol.






B.3 Spec for a reference chamber

Based on results in the literature we can state that *at least* the following meet this specification:

- NE2571 and NE2611
- PTW30010, PTW30012, PTW30013, **PTW31013**
- Exradin A12, A12S, A19, **A18, A1SL**
- IBA FC65-G, FC65-P, FC23-C, **CC25, CC13**
- Capintec PR-06C


- i) majority are 0.6 cm³ 'Farmer-type' chambers
- ii) 5 scanning chambers, NO microchambers
- iii) A-150 chambers **explicitly excluded**
- iv) Parallel-plate chambers also excluded

B.4 Polarizing voltage

Based on results in the literature we can state the following:

- Not all chambers follow standard 'Boag' theory
- Manufacturers' statements on voltage limits need verifying (**at least** for chamber types, if not individual chambers)
- Going to a higher polarizing voltage can lead to a larger uncertainty in the measurement
- Recombination can be a function of the sign of the charge collected
- **accurate to within 0.5%. It is the responsibility of the user to confirm this** by comparing the results to those of a calibrated reference chamber for which data are given in the protocol




B.5 Measurement of P_{pol}

The polarity correction should be measured for any new chamber and beam combination. **It doesn't take very long.**

The measurement of P_{pol} is a very simple QA check of the chamber/electrometer system:


- i) it confirms that the polarizing voltage is applied correctly between the chamber's electrodes,
- ii) chamber-to-chamber variations in P_{pol} tend to be small – any deviation from published values may indicate non-standard behaviour
- iii) any change in P_{pol} with time indicates a possible change in chamber response.



B.7 Use of small-volume chambers

- Very small chambers (volumes < 0.05 cm³) are not recommended for reference dosimetry. **They do not meet the specification for a reference dosimeter.**
- Issues include: anomalous recombination behaviour, large polarity effect, long settling times, leakage, cable currents.
- These can also impact relative dosimetry measurements (such as measurement of depth-dose curves or beam profiles)
- Careful characterization of such chambers is recommended before use in **any situation.**



accurate to within 0.5%. It is the responsibility of the user to confirm this by comparing the results to those of a calibrated reference chamber for which data are given in the protocol



Part 2 – work required


$$k_Q = P_{gr} k_{ecal} k'_{R50}$$

- P_{gr} - need to revisit in light of new effective point of measurement data for photon beams.
- k_{ecal} – improved value available in literature for NACP-02 chamber, new values required for new chamber types (Exradin A10, IBA PPC-05, PTW34045).
- k'_{R50} – TG-51 assumes that well-guarded parallel-plate chambers have a unity perturbation correction. Multiple publications now show this is not the case – new data required.






1. Ion-chamber perturbation corrections


Monte Carlo investigations



References: Verhaegen et al (PMB, 2006), Buckley and Rogers (Med. Phys., 2006), Araki (Med. Phys., 2008)



Measurement of k'_{R50}



Reference: Cojocaru et al, (IAEA Symposium, 2010)


MC data from: Zink and Wulff (PMB, 2010)

Early days – more work required to develop electron beam primary standards and evaluate suitable reference chambers

Conclusions

- **There is still interesting work to be done in the field of reference dosimetry**
- **TG-51 is looking (pretty) good as it moves into a second decade**
- **Some changes are required – both photons and electrons**
- **Keep an eye out for published Addenda**



National Research Council Canada / Conseil National de Recherches Canada



Thank you

National Research Council Canada / Conseil National de Recherches Canada

Canada 20
