2D Digital Mammography-An update on vendor-recommended QC tests

D. R. Jacobson, Ph.D. Asst. Professor of Radiology Medical College of Wisconsin Milwaukee, WI 53226

We will consider:

- 1. Benefit of screening mammography
- 2. Importance of good image quality in mammography
- 3.Value of QC in mammography
- 4.Current QC program requirements for FFDM
- 5. Potential changes for the future

Age-adjusted Cancer Death Rates for Females in the US, 1930-2009



American Cancer Society. Cancer Facts & Figures 2013. Atlanta: American Cancer Society; 2013.

Age-adjusted Cancer Death Rates for Females in the US, 1930-2009



American Cancer Society. Cancer Facts & Figures 2013. Atlanta: American Cancer Society; 2013.

Survival is very good if the cancer is found early

5-year survival rates (%):
98 if cancer is local
84 if regional
24 if metastasized

American Cancer Society. Cancer Facts & Figures 2013. Atlanta: American Cancer Society; 2013.

Why do we do mammography? What is the purpose of mammography?

early and accurate detection of breast cancer

with the minimum required radiation dose

How is breast cancer detected in a mammogram?

How is breast cancer detected in a mammogram?

What is mammography image quality?

What is the benefit of QC in mammography? How does it assure good image quality? How does it improve breast cancer detection?

Quality control should not be thought of as an end in itself

Pathology to be seen:

high density mass

shape and margin are important



Pathology to be seen:

micro-calcifications

morphology and distribution are important



Pathology to be seen:

architectural distortion

deviation from normal parenchymal pattern



How is breast cancer detected in a mammogram?



Tabar L, Tot T, Dean PB. Breast cancer: the art and science of early detection with mammography. Stuttgart, Germany, Thieme Verlag, 2005

How is breast cancer detected in a mammogram?

benign stellate lesions 7%



"...perception of the stellate lesion—while they are small [is] the number one task of mammography" L. Tabar

Tabar L, Tot T, Dean PB. Breast cancer: the art and science of early detection with mammography. Stuttgart, Germany, Thieme Verlag, 2005

Mammography QC- a little history:

- 1985 NEXT (FDA)
- 1986 Galkin et al

Wide variability in image quality and radiation dose in radiography

1986 ACS establishes Breast Cancer **Awareness Screening Program ACS-ACR** collaboration sets standard for quality mammography **1987- ACR launches voluntary** accreditation program including mandatory QC testing

A little more history

- 1987: ACR initiates voluntary accreditation 1990: 1st ACR QC manuals
- 10/27/92: President Bush signs MQSA bill to take effect 10/1/94 (21CFR Part 900)
- 1992: 2nd ACR manual published
- 10/23/93: Interim MQSA regulations published 10/28/97: Final MQSA regulations published to take effect 4/28/99 (a few effective 10/28/02) 1999: New ACR manual to work with MQSA

The ACR QC program



• FDA certification and ACR accreditation Mammography Quality Standards Act, 1992



Sec. 900.11 Requirements for certification.

(a) *General.* After October 1, 1994, a certificate issued by FDA is required for lawful operation of all mammography facilities subject to the provisions of this subpart. To obtain a certificate from FDA, facilities are required to meet the quality standards in Sec. 900.12 and to be accredited by an approved accreditation body or other entity as designated by FDA.

21CRF 900.12(e)(6):

Quality Control tests — *other modalities*.

For systems with image receptor modalities other than screen-film, the quality assurance program shall be substantially the same as the quality assurance program recommended by the image receptor manufacturer, except that the maximum allowable dose shall not exceed the maximum allowable dose for screen-film systems in paragraph (e)(5)(vi) of this section.

21CRF 900.12(e)(5)(vi):

Dosimetry.

The average glandular dose delivered during a single cranio-caudal view of an FDA-accepted phantom simulating a standard breast shall not exceed 3.0 milligray (mGy) (0.3 rad) per exposure. The dose shall be determined with technique factors and conditions used clinically for a standard breast.

Standard breast is defined as 4.2 cm compressed breast consisting of 50% glandular and 50% adipose tissue

Other modalities (as of August 2013) are:

- Full field digital mammography (FFDM)
 - General Electric 2000D was first 01/28/00
 - Currently 26 different models are approved
- Digital Breast Tomosynthesis (DBT)
 - Hologic Selenia Dimensions DBT 02/11/11

Generic description of selected mammographic QC tests

Medical physicist QC tests for Screen-Film Systems: ACR and MQSA

	PASS/FAIL
1 Mammographic Unit Assembly Evaluation	ACK OUIDES MUSA REUS
2 Collimation Assessment	
Deviation between X ray field and light field <2% of SID	
X ray field does not extend beyond any side of the IP by more than 2% of SID	
Chest wall edge of compression naddle doesn't extend beyond IR by more than 1% of SID	
3 Evaluation of Focal Spot Performance	
Measured performance within acceptable limits for large focal spot	
Measured performance within acceptable limits for small focal spot	
4. Automatic Exposure Control (AEC) System Performance	
Exposure reproducibility is within acceptable limits	
AEC compensation for kVp, breast thickness and image mode is adequate	
AEC density control function is adequate	
5. Uniformity of Screen Speed	
Optical density range is ≤0.30	
6. Artifact Evaluation	
Artifacts were not apparent or not significant	
7. Phantom Image Quality Evaluation	
4 largest fibers, 3 largest speck groups and 3 largest masses are visible	
Phantom image quality scores: Fibers	
Specks	
Masses	
8. kVp Accuracy and Reproducibility	[]
Measured average KVp within ±5% of indicated KVp	
kvp coefficient of variation ≤ 0.02	
9. Beam Quality (Haif-Value Layer) Assessment	
10 Propet Entrenes Expedure Average Clendular Dess and Rediction Output Rate	
Average glandular dose for average breast is below 3 mGy (300 mrad)	
Average glandular dose to a 4.2 cm thick breast on your unit is	
Radiation output rate is >800 mR/sec (7.0 mGv/sec) at 28 kVp with Mo/Mo	
11. Viewbox Luminance and Room Illuminance	·
Mammographic viewbox is capable of a luminance of at least 3000 nit	
Room illuminance (viewbox surface & seen by observer) is 50 lux or less	

Unit Evaluation – flat and parallel compression







.....





Light field – x-ray field – detector alignment



light to x-ray: X and Y sum $\Delta \le 2\%$ SID X-ray to detector: $\Delta \le 2\%$ along any side



Compression paddle to detector alignment



Paddle cannot extend beyond detector by > 1% SID



Focal Spot Performance (System Resolution)



11 lp/mm - length

工业专业

13 lp/mm - width

AEC performance – thickness compensation – density control



Δ O.D. 2-6 cm \leq +/- 0.15



Phantom image quality – MAPP – 4/3/3




kVp accuracy and repeatability accuracy 5% CV ≤ 0.02

Beam quality (HVL) HVL ≥ kVp/100



Entrance exposure, average glandular dose, radiation output rate



Air kerma and mAs $CV \le 0.05$ Output rate $\ge 7.0 \text{ mGy/sec}$ (800 mR/sec) $Dg \le 3.0 \text{ mGy}$ (300 mrad)

Viewbox luminance and viewing conditions (optional under MQSA)

Viewbox luminance $\ge 3000 \text{ cd/m}^2$ Luminance $\le 50 \text{ lux}$







Masking

Currently approved FFDM systems

FFDM and DBT Systems

FDA approved, cleared, or accepted the following FFDM and DBT units for use as indicated by date: Konica-Minolta Siemens Mammomat Inspiration Prime Edition cleared on 6/27/13 Konica Minolta Xpress Digital Mammography Computed Radiography (CR) System on 12/23/11 Agfa Agfa Computed Radiography (CR) Mammography System on 12/22/11 Fuji Aspire Computed Radiography for Mammography (CRM) System on 12/8/11 Giotto Giotto Image 3D-3DL Full-Field Digital Mammography (FFDM) System on 10/27/11 Fuji-DR Fuji Aspire HD Full-Field Digital Mammography (FFDM) System on 9/1/11 GE Senographe Care Full-Field Digital Mammography (FFDM) System on 10/7/11 Planmed Nuance Excel Full-Field Digital Mammography (FFDM) System on 9/23/11 Planmed Planmed Nuance Full-Field Digital Mammography (FFDM) System on 9/23/11 Siemens Mammomat Inspiration Pure Full-Field Digital Mammography (FFDM) System on 8/16/11 Hologic Selenia Encore Full-Field Digital Mammography (FFDM) System on 6/15/11 Philips Philips (Sectra) MicroDose L30 Full-Field Digital Mammography (FFDM) System on 4/28/11 Hologic Selenia Dimensions Digital Breast Tomosynthesis (DBT) System on 2/11/11 Siemens Mammomat Inspiration Full Field Digital Mammography (FFDM) System on 2/11/11 Carestream Carestream Directview Computed Radiography (CR) Mammography System on 11/3/10 Hologic Selenia Dimensions 2D Full Field Digital Mammography (FFDM) System on 2/11/09 Hologic Selenia S Full Field Digital Mammography (FFDM) System on 2/11/09 Siemens Mammomat Novation S Full Field Digital Mammography (FFDM) System on 2/11/09 Hologic Selenia Full Field Digital Mammography (FFDM) System with a Tungsten target in 11/2007 Fuji-CR Fuji Computed Radiography Mammography Suite (FCRMS) on 07/10/06 GE Senographe Essential Full Field Digital Mammography (FFDM) System on 04/11/06 Siemens Siemens Mammomat Novation DR Full Field Digital Mammography (FFDM) System on 08/20/04 GE Senographe DS Full Field Digital Mammography (FFDM) System on 02/19/04 Lorad/Hologic Selenia Full Field Digital Mammography (FFDM) System on 10/2/02 Lorad Lorad Digital Breast Imager Full Field Digital Mammography (FFDM) System on 03/15/02 Fischer Fischer Imaging SenoScan Full Field Digital Mammography (FFDM) System on 09/25/01 GE GE Senographe 2000D Full Field Digital Mammography (FFDM) System on 01/28/00

Common QC tests for FFDM

Mechanical/safety checks Acquisition monitor checks X-ray beam collimation and alignment (dead space) Compression paddle position, flat and parallel Spatial resolution (detector and system: GE) AEC: thickness response, exposure compensation Artifacts/uniformity (flat-field uniformity) Image quality: 4/3/3 or 5/4/4 kVp accuracy and repeatability HVL **Radiation dose** Radiation output rate **SNR/CNR**

Reading workstation Film printer

Common QC tests for FFDM – same as S/F

Mechanical/safety checks Acquisition monitor checks

X-ray beam collimation and alignment (dead space) Compression paddle position, flat and parallel Spatial resolution (detector and system: GE) AEC: thickness response, exposure compensation Artifacts/uniformity (flat-field uniformity) Image quality: 4/3/3 or 5/4/4 kVp accuracy and repeatability HVL Radiation dose Radiation output rate **SNR/CNR**

Reading workstation Film printer

Acquisition Monitor IQ check - SMPTE pattern





Collimation testing without film is a challenge











video ~ 1 sec take snapshot from video







Collimation – sliding paddle

-

э

System Spatial Resolution

Image quality is still assessed with the MAP phantom

Fibers = 4, 5 Specks = 3, 4 Masses = 3, 4



Radiation dose must be < 300 mrad (3 mGy)

SNR and CNR



Current Manufacturer-required QC tests for FFDM

GE Senographe 2000D

1

.





GE Senographe 2000DS



GE Senographe Essential



General Electric

- 1. Flat Field
- 2. Phantom Image Quality

Phantom IQ Test on AWS Phantom IQ Test on Printer

Fibers	Specks	Masses

3. CNR Measurement (NA for DS, Essential or Care if Sub-System MTF test done)

CNR (Required for both new unit Mammography Equipment Evaluations and Annual Surveys)

Change in CNR ≤0.2 (NA for Mammography Equipment Evaluations)

- 4. MTF Measurement (NA for 2000D, DS, Essential or Care if Sub-System MTF test done)
- 5. AOP Mode and SNR
- 6. Collimation Assessment
- Mammo Equipment Evaluation Annual Survey

Ess/Care

DS

2000D

2000D

24 cm x 30.7 cm

19 cm x 23 cm tests

	_

7. Evaluation of Focal Spot Performance (NA for 2000D, DS, Essential or Care if Sub-System MTF test done)

DS

- Sub-System MTF (NA for 2000D if MTF and Focal Spot Performance tests done; NA for DS, Essential or Care if CNR, MTF and Focal Spot Performance tests done)
- Breast Entrance Exposure, Average Glandular Dose and Reproducibility Average glandular dose for average breast is ≤3 mGy (300 mrad)

Exposure reproducibility (CV) for air kerma (R) and mAs is ≤0.05

- 10. Artifact Evaluation and Flat Field Uniformity
- 11. kVp Accuracy and Reproducibility
- 12. Beam Quality Assessment (Half-Value Layer Measurement)
- 13. Radiation Output

Radiation output is ≥800 mR/s

14. Mammographic Unit Assembly Evaluation

Meets requirements for motion of tube-image receptor assembly Meets requirements for compression paddle decompression

15. Review Workstation (RWS) Tests (for all RWS, even if located offsite; NA if only hardcopy read)

	mR/s
--	------

mrad



Contrast to Noise Ratio



- Contrast is difference between means of ROI 1 and 2.
- Noise is the SD of the ROI 2 (background)
- Contrast is measured between background and largest mass

MTF Measurement





Sub-system MTF









Sub-System MTF Test



2.09 and 3.93 lp/mm

5 and 8 lp/mm

Fisher Senoscan



Fischer

- 1. X-Ray Field, Chest Wall Missed Tissue and Light Field Checks
- 2. Compression Paddle Alignment
- 3. kVp Accuracy Test
- 4. Linearity, Reproducibility and Accuracy
- 5. Half-Value Layer and Output
- 6. Dosimetry Average Glandular Dose and Output

Average glandular dose for average breast is ≤3 mGy (300 mrad)

7. Phantom Image Acquisition Test (values required for all tests)

No obvious artifacts		
Background StDev	within +50/-0 ADU counts of baseline (NA for Equipment Evaluations)	
Background mean	within ±100 ADU counts of baseline (NA for Equipment Evaluations)	
ADU level difference	within ±300 ADU counts of baseline (NA for Equipment Evaluations)	

8. Image Quality

Largest 4 fibers, 3 speck groups and 3 masses Phantom IQ Test on Review Work Station

- 9. System Resolution/Scan Speed Uniformity
- 10. Flat Field Test
- 11. Geometric Distortion and Resolution Uniformity
- 12. Automatic Decompression Control
- 13. System Artifacts
- 14. Image Viewing Room Illuminance Test (≤50 lux)
- 15. Review Workstation (RWS) Tests (for all RWS, even if located offsite; NA if only hardcopy read)

Fibers	Specks	Masses

mrad



Hologic/Lorad Selenia



Hologic Selenia Dimensions 2D (3D)

Hologic/Lorad

EE- MP performs RT tests

1.	Mammographic Unit Assembly Evaluation			
2.	Collimation Assessment			
3.	Artifact Evaluation			
4.	kVp Accuracy and Reproducibility			
5.	Beam Quality Assessment - HVL Measurement			
6.	Evaluation of System Resolution			
7.	Automatic Exposure Control (AEC) Function Performance (NA for systems without AEC)			
8.	Breast Entrance Exposure, AEC Reproducibility and Average Glandular Dose			
	Average glandular dose for average breast is ≤3 mGy (300 mrad) mrad			
9.	Radiation Output Rate			
10.	Phantom Image Quality Evaluation			
	Phantom image scores: Fibers Specks Masses			
11.	Signal-To-Noise Ratio and Contrast-To-Noise Ratio Measurements (values required for all tests)			
	SNR (value)			
	CNR (value) (Required for both new unit Mammography Equipment Evaluations and Annual	Surveys)		
	CNR should not vary by more than ±15% (NA for Equipment Evaluation)			
12. Diagnostic Review Workstation (RWS) QC (for all RWS, even if located offsite; NA if only hardcopy read)				
13.	13. DICOM Printer QC (Mammography Equipment Evaluations only)			
14.	4. Detector Flat Field Calibration (Mammography Equipment Evaluations only)			
15.	Compression Thickness Indicator (Mammography Equipment Evaluations only)			
16. Compression (Mammography Equipment Evaluations only)				


Image quality is still assessed with the MAP phantom

Fibers = 5 Specks = 4

Masses = 4





Siemens Mammomat Novation





Siemens Inspiration

Siemens

1.	Site Audit/Evaluation of Technologist QC Program	
•		

- 2. Mechanical Inspection
- 3. Acquisition Workstation Monitor Check
- 4. Detector Uniformity
- 5. Artifact Detection
- 6. Collimation, Dead Space & Compression Paddle Position
- 7. AEC Thickness Tracking
- 8. Spatial Resolution
- 9. SNR, CNR and AEC Repeatability

Measured values: SNR CNR CNR CV for mAs and entrance air kerma <5%

Max deviation of mean pixel values and SNR within ±15% of mean for measurements

10. Image Quality

Largest 5 fibers, 4 speck groups and 4 masses visible*

(*largest 4 fibers, 3 speck groups and 3 masses acceptable if spatial resolution and CNR pass)

Phantom image scores: Fibers Specks Masses

11. Radiation Dose

Average glandular dose for average breast is <3 mGy (300 mrad)

- 12. HVL and Radiation Output
- 13. Tube Voltage Measurement & Reproducibility
- 14. Film Printer Check

15. Review Workstation (RWS) Tests (for all RWS, even if located offsite; NA if only hardcopy read)



I			
I			
I			
I			











- Contrast is difference between means of ROI 1 and 2.
 Noise is the SD of the
- Noise is the SD of the ROI 2 (background)

Image quality is still assessed with the MAP phantom

Fibers = 5 Specks = 4

Masses = 4



Radiation dose **must not** exceed 3 mGy Radiation dose **should not** exceed 2 mGy

Pre-requisites to calculate the mean glandular dose

Use formulae:

D=Kgcs

[Dance et al 2000]

- D Mean glandular dose
- K Entrance air kerma
- g g-factor for breasts simulated with PMMA
- c c-factor for breasts simulated with PMMA
- s s-factor for clinically used spectra

Where *K* is the incident air kerma (without backscatter) calculated at the upper surface of the PMMA. The factor *g*, corresponds to a glandularity of 50%, and is derived from the values calculated by Dance et al 2000 and is shown below for a range of HVL. The *c*-factor corrects for the difference in composition of typical breasts from 50% glandularity [Dance et al 2000] and is given here for typical breasts in the age range 50 to 64. Note that the *c* and *g*-factors applied are those for the corresponding thickness of typical breast rather than the thickness of PMMA block used. Where necessary interpolation may be made for different values of HVL. The factor *s* shown in the second table corrects for any difference due to the choice of X-ray spectrum (Dance et al 2000).

	PMMA thickness	Equivalent breast				HVL (r	nm Al)			
	(mm) thickne (mm)	(mm)	0.25	0.30	0.35	0.40	0.45	0.50	0.55	0.60
g-factor	45	53	0.130	0.155	0.177	0.198	0,220	0.245	0,272	0,295
c-factor	45	53	-	1.109	1.105	1.102	1.099	1.096	1.091	1.088
g-factor	50	60	0.112	0.135	0.154	0.172	0.192	0.214	0.236	0.261
c-factor	50	60	-	1.164	1.160	1.151	1.150	1.144	1.139	1.134

	Mo/Mo	Mo/Rh	W/Rh
s-factor	1.000	1.017	1.042



Fuji ClearView CR*m*





Fuji CR

	_	
1.	S Value Confirmation (≤120 ±20% [96 ≤ corrected S value ≤ 144])	
2.	System Resolution (8 lp/mm <u>+</u> 2 lp/mm in both directions)	
3.	CR Reader Scanner Performance	
4.	Mammography Unit Assembly Evaluation	
5.	Collimation Assessment Test date if different from above:	
	Chest wall edge of X-ray field extends to edge of IR	
	Deviation between X-ray field and light field $\leq 2\%$ of SID	
	X-ray field does not extend beyond any side of the IR by more than 2% of SID	
	Paddle chest wall edge not beyond IR by more than 1% of SID or appear on the image	
6.	Automatic Exposure Control (AEC) System Performance Assessment	
	AEC density control function meets Fuji performance criteria	
	Reproducibility (CV) for either exposure or mAs is ≤0.05	
	Image mode tracking meets Fuji performance criteria	
	CNR per object thickness meets Fuji performance criteria	
7.	System Artifact Evaluation	
8.	Phantom Image Quality Evaluation Fibers Specks Masses (at least 4 fibers, 3 speck groups	s & 3 masses)
	Phantom IQ (printed images)	
	Phantom IQ (softcopy)	
	Other tests meet Fuji performance criteria (mAs, OD & DD for hardcopy, S value for soft copy)	
9.	Dynamic Range	
10.	0. Primary Erasure (Additive and Multiplicative Lag Effects)	
11.	1. Inter-Plate Consistency (variation of mAs within ±10%; SNR within ±15%)	
12.	2. kVp Accuracy and Reproducibility Test date if different from above:	
	Measured average kVp within <u>+</u> 5% of indicated kVp	
	kVp coefficient of variation <0.02	
13.	3. Dose (average glandular dose for average breast is ≤3 mGy [300 mrad]) mrad	
14.	Beam Quality Assessment & HVL Measurement Test date if different from above: HVL ≥kVp/100 mm AI	
15.	5. Radiation Output Test date if different from above:	
	Radiation output rate is <a>800 mR/sec (7.0 mGy/sec) at 28 kVp with Mo/Mo	
16.	5. Viewing and Viewing Conditions Test date if different from above:	
	Room illuminance <20 lux or as recommended by the monitor manufacturer	
17.	7. Review Workstation (RWS) Tests (for all RWS, even if located offsite; NA if only hardcopy read)	
18.	3. Film Printer Tests (for all printers used for mammography, even if located offsite)	

At 20 mR, for 25 kVp and Mo/Mo:

S= 120 +/- 20% (96-144)

Expose – read should be 5-10 minutes















Carestream DirectView

Carestream

	= $=$ $=$ $=$ $=$ $=$ $=$ $=$ $=$ $=$	
1.	Cassette Exposure Response Test NOT target = 1000 x 10g ₁₀ (IIIX) + 1000	
2.	Cassette/Phosphor Screen Artifact Test	
3.	Erased Screen Test	
4.	Scanner Uniformity Test	
5.	Scanner Response Linearity Test	
6.	Spatial Frequency Response Test	
7.	Geometric Accuracy Test	
8.	Image Plate Fog Test	
9.	AEC System Performance/Constancy Test	
	AEC Function with Breast Thickness	
	AEC Density Control Function	
	CNR Tracking with Breast Thickness	
10.	Mammography Unit Assembly Evaluation	
11.	Mammographic Unit Collimation Assessment Test date if different from above:	
	X-ray Field to Light Field Alignment	
	X-ray Field to Image Plate Alignment	
	Alignment of Chest Wall Edge of Compression Paddle to Image Plate	
12.	Beam Quality and Half-Value Layer Test date if different from above:	
13.	kVp Accuracy and Reproducibility Test date if different from above:	
	Measured average kVp within ±5% of indicated kVp	
	kVp coefficient of variation ≤0.02	
14.	Breast Entrance Exposure, Dose and Radiation Output	
	Average glandular dose for average breast is below 3 mGy (300 mrad)	
	Average glandular dose to a 4.2-cm-thick breast on your unit is	
	Radiation output rate is >800 mR/sec (7.0 mGy/sec) at 28 kVp with Mo/Mo	
15.	Phantom Image Quality Evaluation Fibers Specks Masses (at least 4 fibers, 3 speck gr	oups & 3 masses)
	Phantom IQ (printed images)	
	Phantom IQ (softcopy)	
16.	Review Workstation (RWS) Tests (for all RWS, even if located offsite; NA if only hardcopy read)	
17.	Film Printer Test (for all printers, even if located offsite)	

QC Manual – Important Points



Calibration = 5 minutes TQT = 5 minutes QC testing = 5 minutes

Unrestricted Internal Use © 2011, Carestream Health





Sectra MicroDose Mammography



System uses scanning multi-slot geometry

X-ray exposure is pulsed: 10-50 msec, 5-30 pulses/sec.

mAs \approx 1% conventional mAs

Assure that exposure and kVp meters will work with this system

Place Pb aperture on paddle for HVL

Philips (Sectra)





The average glandular dose (AGD) is calculated by

AGD = ESAK'g'c's,

where g and c are functions of the breast thickness and HVL and s is a function of breast or PMMA thickness. The g-, c- and s-tables can be interpolated. The entrance surface air kerma (ESAK) should be calculated at the entrance surface of the breast.

Table 11 g-factor as a function of breast thickness and HVL [10].Table 12 c-factors for average breasts for women 50 to 64 [10].Table 13 s-factors

 $Dg \le 1 mGy (100mrad)$



Contrast-to-noise ratio measured using the Daily QC Phantom





Planmed Nuance

Planmed

- 1. Monitor Cleaning (AWS and RWS) (Mammography Equipment Evaluation only)
- 2. Monitor Quality (AWS) TG-18 QC test phantom (Mammography Equipment Evaluation only)
- 3. Phantom Image Quality (AWS and RWS) (Mammography Equipment Evaluation only)

Phantom IQ Test on AWS Phantom IQ Test on RWS

.

- 4. Viewbox and Viewing Conditions (Mammography Equipment Evaluation only)
- 5. Signal Homogeneity (Mammography Equipment Evaluation only)
- 6. Uncorrected Defective Elements (DEL) (Mammography Equipment Evaluation only)
- 7. Large Focus Calibration (Mammography Equipment Evaluation only)
- 8. Small Focus Calibration (for mags only) (Mammography Equipment Evaluation only)
- 9. Signal-to-Noise (SNR) (Mammography Equipment Evaluation only)
- 10. Contrast-to-Noise Ratio (CNR) (Mammography Equipment Evaluation only)
- 11. Visual Checklist (Mammography Equipment Evaluation only)
- 12. Repeat Analysis (Mammography Equipment Evaluation only)
- 13. Defect Acceptance Test (Mammography Equipment Evaluation only)
- 14. System Fault Report (Mammography Equipment Evaluation only)
- 15. Review Workstation QC-Overall (Mammography Equipment Evaluation only)
- 16. Film Printer QC (Mammography Equipment Evaluation only)
- 17. Ghosting
- 18. Modulation Transfer Function (MTF)
- 19. Linearity/Noise Linearity
- 20. AEC
- 21. Compression Force
- 22. Mammographic Unit Assembly
- 23. Beam Quality Assessment HVL Measurement
- 24. Breast Entrance Exposure and Average Glandular Dose

Average glandular dose for average breast is ≤3 mGy (300 mrad)

mrad

FUJI Aspire

Fuji Aspire HD

1. 1Shot PhantomM

Missed Tissue on Chest Wall Edge CNR /Nammography Equipment Evaluation only? 1 Shot Phantom Sensitivity Constancy /Nammography Equipment Evaluation only? Geometric Distortion /Nammography Equipment Evaluation only? System Artifact Evaluation /Nammography Equipment Evaluation only? Uniformity /Nammography Equipment Evaluation only? Dynamic Range (Nammography Equipment Evaluation only? Spatial Resolution (SR) /Nammography Equipment Evaluation only? Low Contrast Detectability (LCD) /Nammography Equipment Evaluation only? Linearity/Beam Quality Constancy (Nammography Equipment Evaluation only?

2. ACR Phantom (Mammography Equipment Evaluation only)

Phantom IQ Test on AWS Phantom IQ Test on RWS

Fibers	Specks	Masses

- 3. Image Basic Test
- 4. Compression Device Confirmation
- 5. Viewbox Maintenance
- 6. Monitor Quality Control (Secondary/AWS)
- 7. Additive Lag Effects
- 8. Multiplicative Lag Effects (Ghost)
- 9. Visual and Functional Test
- 10. Spatial Resolution (Magnification)
- 11. kVp Accuracy and Reproducibility
- 12. Half Value Layer (HVL)
- 13. Collimation Assessment
- 14. Radiation Output
- 15. AEC Reproducibility
- 16. CNR Mode 1
- 17. AGD Mode 1
- 18. AGD-ACR Phantom

Average glandular dose for average breast is ≤3 mGy (300 mrad)

- 19. Review Workstation QC-Overall
- 20. Film Printer QC-Overall

	_
	٦
	1
	J
	1
	1
	1
	4
	1
	J
	J
	1
	-
	1
	1
	J
	1
	J
	1
	÷
	J
	1
	_
	1
	J
	1
	J
	1
	1
	1
	-

Gioto Image

Giotto

4	Collimation Assessment	•
· · ·		
2.	Artiact Evaluation	
э.	Automatic Exposure Control (AEC), Signal-To-Noise Ratio (SNR) & Contrast-To-Noise Ratio (Ch	
	SNR (value) ≥60	
	CNR (value) (Required for both new unit Mammography Equipment Evaluations and Annual S	Surveys)
	CNR should not vary by more than ±20% (NA for Equipment Evaluation)	
4.	AEC Reproducibility	
5.	ACR Phantom Image Quality	,
	Phantom image scores: Fibers Specks Masses	
6.	Ghost Factor	
7.	Inactive Border at Chest Wall Edge	
8.	Flat Field Homogeneity	
9	Detector Response Function and Noise Evaluation	
10	Spatial Resolution	
11	kVn Accuracy and Reproducibility	
42	Tube Output	
12.		
13.		
14.	. Beam Quality (HVL)	
15.	5. Mean Glandular Dose (MGD)	
	Average glandular dose for average breast is ≤3 mGy (300 mrad) mrad	
16.	6. Viewbox Luminance	
17.	. Diagnostic Review Workstation (RWS) QC (for all RWS, even if located offsite; NA if only hardcopy read)	
18.	8. Film Printer QC	



Konica-Minolta

Xpress CR - Contact Mammography Upgrade



Konica Minolta



QC PHANTOMS AND AUTOMATED ANALYSIS PROGRAMS: GE

flat field test AOP and SNR IQST (Image Quality Signature Test) Phantom- MTF and CNR



Flat Field Tests AWS Screen Results

Image Qual	ity Test Resu	lts 2000-12-29,
Test	Action Level	Measurement
Brightness Uniformity	= 10%</td <td>1.82</td>	1.82
High Frenquency modulation	= 0.8%</td <td>0.49</td>	0.49
Bad Pixel	<100	0.00
Bad ROI	None	0.00
SNR Uniformity	= 40%</td <td>23.03</td>	23.03

1. Flat Field and In	(Weekly technologist test.)					
a. Fl	at Field: 2.5 cm thick uniform	; Mo/Mo. Ph	antom evalua	nted on AWS		
	Test		Status	Objects	Observed	Status
Bi	rightness Uniformity < 10%	1.82	Pass	Fibers	5.5	Pass
F	High Freq Modulation < 0.8		Pass	Specks	4	Pass
	Bad Pixel < 100		Pass	Masses	4	Pass
	Bad ROI = 0		Pass			
	SNR Uniformity < 40%		Pass			
b. Pl	b. Phantom Image Quality (Print, AWS and RWS).					
RMI	RMI Model156 Phantom: 26 kVp:125 mAs, Mo/Mo.					

Image	Quality	Test	Results	2009-0	5-19	,09:31	:20,0	DT
-------	---------	------	---------	--------	------	--------	-------	----

N.

Test	Measurement	LSL	USL	Status
Brightness Non Uniformity	1.66	N/A	10.00	PASS
High Frequency Modulation	0.49	N/A	0.80	PASS
Bad Pixel	0.00	N/A	100.00	PASS
Bad ROI	0.00	N/A	0.00	PASS
SNR Non Uniformity	28.67	N/A	40.00	PASS

X

OK







DS IQ Signature Test automates SNR and MTF







Put 25mm acrylic plates on the bucky Align the longest side of the plates with the chestwall edge of the bucky. Center the plates left-right. Apply a compression force of 5 daN. Perform 1 exposure.

QC PHANTOMS AND AUTOMATED ANALYSIS PROGRAMS: Hologic

Auto SNR and CNR

Contrast to Noise Ratio





QC PHANTOMS AND AUTOMATED ANALYSIS PROGRAMS: Fuji CR

Auto SNR and CNR





QC calculation tool - offline

age select		_	-		The second second	and the second second		
mage for calculation	1D	-	Patientname	MenuNane	Imageleate	Image lime	Hostnan	Bystem Resolution
		34	DRIestiu	PHYSICS	20080318	200405.828	CIERNAM	Stran Direstion
		33		PHYSICS	20080318	193857.546	CLRVW	
		32		PHYSICS .	20080318	181731.687	CLRVW	Contrary Discontrations
		31		PHYSICS	20080318	175628 140	GERVW	But-scan Direction
		30	DJTest6	PHYSICS	20080318	174426 343	GLRVW	
		29	DJTest5	PHYSICS I_	20080318	150324 562	CLRVW	
		28	DJTest4	PHYSICS	20080318	134524.234	CLRVW_	CNR
		27	DJTest3	PHYSICS_	20080318	130436 078	CLRVW	CHIT
	lat	26	DiTect2	PHYSICS	20080318	120643/156	ICLEVANCE.	
23								
niformly exposed image	1D		Patientname	MenuName	ImageDate	ImageTime	Hostnan	Standard Deviation
		34	DRTest10	PHYSICS	20080318	200405.828	CLRVW	
		33		PHYSICS	20080318	193857 546	CLRVW	-
		32		PHYSICS	20080318	181731.687	CLRVW	SNR
		31		PHYSICS	20080318	175628 140	CLRVW	
		30	DJTest6	PHYSICS	20080318	174426 343	CLRVW	
		29	DJTest5	PHYSICS 1_	20080318	150324 562	CLRVW	Homogeneity
		28	DJTest4	PHYSICS	20080318	134524.234	CLRVW	Thomagaining
		27	DJTest3	PHYSICS	20080318	130436.078	CLRVW	
	NI.	26	D.ITest2	PHYSICS	20080318	120643.156	CLENN	
	101							
Calculation								
Calculation							_	
Mean	1	SN	R					
SNR	1.412	3	71.712					

QC PHANTOMS AND AUTOMATED ANALYSIS PROGRAMS: Fuji DR

Fuji One Shot phantom and auto QC program

Fuji 1 Shot Phantom M Plus weekly constancy test tool



9.1 Specification Outline of 1Shot Phantom M Plus

Test Contents	Applicable Phantom	Outline of Calculation Method
Missed tissue on chest wall edge	Bar patterns (1 mm spacing)	The amount of missed tissue on the chest wall-side edge is calculated by using coordinate information of the slit.
CNR	Al (0.2 mm)	CNR is calculated according to the definitional equation described in IEC 61223-3-2.
1Shot Phantom sensitivity constancy	Uniform exposure region at 60 mm from the chest wall- side edge	A calculation area (20 x 20 mm) is set at the lateral center and also at 60 mm from the chest wall-side edge. Based on the pixel value in this calculation area, the value inversely proportional to the air kerma is calculated. Note: Although this value is calculated in the same way as S value, this value and S value need to be discriminated.
Geometric distortion	Lines indicated with triangle markers (100 mm spacing)	The distance between the two lines is calculated.
System artifact evaluation	Uniform exposure region made of metal and plastic (The transmission factor becomes equivalent to 40-mm PMMA phantom at 28kVp, Mo/Mo.)	N/A
Uniformity (Uniform Exposure Region)	Uniform exposure region at 60 mm from the chest wall- side edge and uniform exposure regions at four corners of the phantom	A calculation area (10 x 10 mm) is set at the lateral center and also at 60 mm from the chest wall. Based on this calculation region, the pixel value and relative SNR value are calculated in the calculation regions (10 x 10 mm each) at four corners of the phantom.
Dynamic range	The lightest region of the step wedge	The pixel value in the highest density region of the step wedge is calculated. According to IEC 61223-3-2, calculation is executed only in the highest density region.
Spatial Resolution (SR)	Bar patterns slanted at 45 degrees (2, 4 and 8 cycles/ mm)	SCTF is calculated according to the definitional equation described in IEC 61223-3-2.
Low Contrast Detectability (LCD)	PMMA phantom (φ 2 mm, contrast 1.4%)	A cross-correlation factor is calculated based on the reference image (ideal image).
Linearity/Beam quality constancy	Five step-wedge patterns forming a density region in the range of the center density plus or minus about one digit	The pixel value is calculated in each step of the step wedge and then the difference between the adjacent steps is calculated.











QC PHANTOMS AND AUTOMATED ANALYSIS PROGRAMS: Carestream

CR Mammo TQT

Manual QC Procedure vs. CR Mammo TQT

The DIRECTVIEW Total Quality Tool (TQT) enables you to perform objective image tests and QC measurements with the same interface used for examinations.



Manual QC Procedure vs CR Mammo TQT

- Manual QC procedure satisfies 21 CFR 900, TQT does not
- TQT is a scanner test only
- Performed by Service upon install
- Recommended for sites to perform twice a year
- Some states make it a yearly requirement
- Does not change the QC testing that has to be performed





Potential future changes

- ACR FFDM QC program
- Universal phantom
- Alternative standard (MQSA)

New ACR accreditation phantom (proposed)



Who determines the QC requirements for a new mammography modality?

- 1. FDA
- 2. ACR
- 3. Imaging system manufacturer
- 4. On-site physicist

5. Image receptor manufacturer

Which of the following are new modalities- requiring 8 hours of initial training before use?

- Screen/film and stereo mammography
 Full-field digital mammography and digital breast tomosynthesis
- 3. Stereotactic biopsy
- 4. Digital stereotactic biopsy
- 5. Automated breast ultrasound

<u>www.fda.gov/Radiation-EmittingProducts/MammographyQualityStandardsActandProgram</u> /Guidance/PolicyGuidanceHelpSystem/ucm136925.htm <u>www.fda.gov/Radiation-EmittingProducts/MammographyQualityStandardsActandProgram</u> /FacilityCertificationandInspection/ucm243765.htm

The 5 year survival for a minimal breast cancer is

- Less then 50%
 60-70%
 70.80%
- 3.70-80%
- 4.80-90%
- 5. Greater than 90%

American Cancer Society. *Cancer Facts & Figures 2013*. Atlanta: American Cancer Society; 2013

The most reliable indication of breast cancer in a mammogram is

- 1. Architectural distortion
- 2. Clustered calcifications
- 3. Linear and branching calcifications
- 4. Spiculated mass
- 5. Skin thickening

Tabar L, Tot T, Dean PB. Breast cancer: the art and science of early detection with mammography. Stuttgart, Germany, Thieme Verlag, 2005

What performance parameter(s) does MQSA specify for FFDM quality assurance?

- 1. Phantom dose < 3.0 mGy for one CC view
- 2. Phantom dose < 2.0 mGy for one CC view Image
- 3. Image quality must exceed S/F IQ

4. Dose can't exceed max. allowed S/F dose

5. Patient dose \leq 3.0 mGy for one CC view